

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062518\
 Data File : VN049543.D
 Acq On : 26 Jun 2018 9:09
 Operator : MD\SY
 Sample : J3684-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1-S

Quant Time: Jun 26 13:24:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	1212954	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	8.58	114	1927461	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	1636714	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	729379	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	795862	49.47	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	98.94%	
35) Dibromofluoromethane	7.58	113	720614	46.21	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	92.42%	
50) Toluene-d8	10.08	98	2619862	44.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.24%	
62) 4-Bromofluorobenzene	12.40	95	829445	42.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.50%	
Target Compounds						
16) Acetone	3.81	43	258312	66.13	ug/l	90
20) Methylene Chloride	4.54	84	502479	30.80	ug/l	92
67) Ethyl Benzene	11.51	91	84453	1.21	ug/l	97
68) m/p-Xylenes	11.61	106	91503	3.49	ug/l	98
69) o-Xylene	11.95	106	29001	1.15	ug/l	89
84) 1,2,4-Trimethylbenzene	13.04	105	134066	2.68	ug/l	96
95) Naphthalene	15.13	128	527896	26.42	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062518\
Data File : VN049543.D
Acq On : 26 Jun 2018 9:09
Operator : MD\SY
Sample : J3684-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1-S

Quant Time: Jun 26 13:24:34 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

