

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062519\
 Data File : VN056484.D
 Acq On : 25 Jun 2019 14:18
 Operator : JC/SP
 Sample : K3510-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-04-062419


Quant Time: Jun 26 03:42:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 03:45:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	354511	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	605299	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	603732	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	169019	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	205700	52.35	ug/l	0.00
Spiked Amount			Recovery	=	104.70%	
35) Dibromofluoromethane	7.59	113	176508	46.74	ug/l	0.00
Spiked Amount			Recovery	=	93.48%	
50) Toluene-d8	10.09	98	757896	51.58	ug/l	0.00
Spiked Amount			Recovery	=	103.16%	
62) 4-Bromofluorobenzene	12.40	95	241871	49.38	ug/l	0.00
Spiked Amount			Recovery	=	98.76%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	12039	6.116	ug/l	100
18) Methyl Acetate	4.34	43	9642	1.662	ug/l	95
20) Methylene Chloride	4.55	84	27541	6.650	ug/l	98
29) Tetrahydrofuran	7.24	42	1436	0.902	ug/l #	58
30) Chloroform	7.37	83	2815	0.454	ug/l	92
31) Cyclohexane	7.67	56	5605	0.924	ug/l #	1
36) 1,1-Dichloropropene	7.66	75	22707	4.262	ug/l #	49
38) Carbon Tetrachloride	7.66	117	31017	6.657	ug/l #	14
43) Isopropyl Acetate	8.28	43	121435	15.041	ug/l #	54
44) Trichloroethene	8.84	130	2403	0.544	ug/l	81
45) 1,2-Dichloropropane	9.27	63	1884	0.430	ug/l #	43
48) Methyl methacrylate	9.18	41	8789	2.172	ug/l #	12
49) 1,4-Dioxane	9.26	88	30	0.357	ug/l #	1
51) 4-Methyl-2-Pentanone	10.02	43	7921	1.520	ug/l #	35
54) cis-1,3-Dichloropropene	9.90	75	61393	10.483	ug/l #	62
57) 1,3-Dichloropropane	10.76	76	12136	1.794	ug/l #	42
59) 2-Hexanone	10.76	43	160354	41.870	ug/l #	51
71) Bromoform	12.40	173	1612	3.599	ug/l #	1
73) Isopropylbenzene	12.40	105	148	Below Cal	#	48
74) N-amyl acetate	11.88	43	12983	2.414	ug/l #	50
76) 1,2,3-Trichloropropane	12.40	75	110921	31.650	ug/l #	100
79) 2-Chlorotoluene	12.73	91	272	Below Cal	#	40
80) 1,3,5-Trimethylbenzene	12.73	105	3814	Below Cal		94
81) trans-1,4-Dichloro-2-buten	12.40	75	110921	112.488	ug/l #	8
83) tert-Butylbenzene	12.73	119	415	Below Cal	#	83
95) Naphthalene	15.13	128	280	5.277	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. 0.00 min

Lab File: VN056484.D

Acq: 25 Jun 2019 14:18

Instrument :

MSVOA_N

ClientSampleId :


TR-04-062419

Tot Ion:168 Resp: 354511

Ion Ratio Lower Upper

168 100

99 50.1 40.6 61.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN

7.67



150000

100000

50000

0

Time--> 7.50 7.60 7.70 7.80



#16

Acetone

Concen: 6.116 ug/l

RT: 3.82 min Scan# 637

Delta R.T. 0.00 min

Lab File: VN056484.D


Acq: 25 Jun 2019 14:18

Tgt Ion: 43 Resp: 12039

Ion Ratio Lower Upper

43 100

58 33.0 26.4 39.6



Abundance Ion 43.00 (42.70 to 43.70): VN

Ion 58.00 (57.70 to 58.70): VN

3.82

4000


3000


2000

1000

0

Time--> 3.80 3.90









#18
Methyl Acetate
Concen: 1.662 ug/l
RT: 4.34 min Scan# 797
Delta R.T. 0.01 min
Lab File: VN056484.D
Acq: 25 Jun 2019 14:18




Instrument :
MSVOA_N
ClientSampled :
TR-04-062419


Tot Ion: 43 Resp: 9642
Ion Ratio Lower Upper
43 100
74 21.6 19.2 28.8



#20
Methylene Chloride
Concen: 6.650 ug/l
RT: 4.55 min Scan# 863
Delta R.T. 0.00 min
Lab File: VN056484.D
Acq: 25 Jun 2019 14:18

Tgt Ion: 84 Resp: 27541
Ion Ratio Lower Upper
84 100
49 119.6 96.4 144.6
51 37.8 28.6 43.0
86 66.3 51.4 77.2





#29

Tetrahydrofuran

Concen: 0.902 ug/l

RT: 7.24 min Scan# 1699

Delta R.T. 0.02 min

Lab File: VN056484.D

Acq: 25 Jun 2019 14:18

Instrument :

MSVOA_N

ClientSampled :

TR-04-062419


Tot Ion: 42 Resp: 1436

Ion Ratio Lower Upper

42 100

72 14.1 34.7 52.1#


71 16.9 33.0 49.6#





Abundance Ion 42.00 (41.70 to 42.70): VN0

Ion 72.00 (71.70 to 72.70): VN0

Ion 71.00 (70.70 to 71.70): VN0



Time--> 7.15 7.20 7.25 7.30



#30

Chloroform

Concen: 0.454 ug/l

RT: 7.37 min Scan# 1741

Delta R.T. 0.00 min

Lab File: VN056484.D


Acq: 25 Jun 2019 14:18

Tgt Ion: 83 Resp: 2815

Ion Ratio Lower Upper


83 100

85 71.6 52.3 78.5





Abundance Ion 82.90 (82.60 to 83.60): VN0

Ion 84.90 (84.60 to 85.60): VN0



Time--> 7.30 7.35 7.40









#31
Cyclohexane
Concen: 0.924 ug/l
RT: 7.67 min Scan# 1833
Delta R.T. 0.02 min
Lab File: VN056484.D
Acq: 25 Jun 2019 14:18

Instrument :
MSVOA_N
ClientSampleId :
TR-04-062419

Tot Ion: 56 Resp: 5605
Ion Ratio Lower Upper
56 100
69 202.8 25.8 38.6#
84 152.9 69.4 104.0#




Abundance Ion 56.00 (55.70 to 56.70): VN0
Ion 69.00 (68.70 to 69.70): VN0
Ion 84.00 (83.70 to 84.70): VN0






#33
1,2-Dichloroethane-d4
Concen: 52.353 ug/l
RT: 8.03 min Scan# 1945
Delta R.T. 0.00 min
Lab File: VN056484.D
Acq: 25 Jun 2019 14:18

Tgt Ion: 65 Resp: 205700
Ion Ratio Lower Upper
65 100
67 54.2 0.0 109.0



Abundance Ion 64.90 (64.60 to 65.60): VN0
Ion 66.90 (66.60 to 67.60): VN0









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.59 min Scan# 2119
 Delta R.T. 0.00 min
 Lab File: VN056484.D
 Acq: 25 Jun 2019 14:18

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-04-062419

Tot Ion:114 Resp: 605299
 Ion Ratio Lower Upper
 114 100
 63 19.5 0.0 39.2
 88 15.2 0.0 31.4




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VN
 Ion 87.90 (87.60 to 88.60): VN






#35
 Dibromofluoromethane
 Concen: 46.744 ug/l
 RT: 7.59 min Scan# 1810
 Delta R.T. 0.00 min
 Lab File: VN056484.D
 Acq: 25 Jun 2019 14:18

Tgt Ion:113 Resp: 176508
 Ion Ratio Lower Upper
 113 100
 111 101.8 82.9 124.3
 192 21.8 17.7 26.5



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V









#36
 1,1-Dichloropropene
 Concen: 4.262 ug/l
 RT: 7.66 min Scan# 1832
 Delta R.T. -0.13 min
 Lab File: VN056484.D
 Acq: 25 Jun 2019 14:18




Instrument :
 MSVOA_N
 ClientSampleId :
 TR-04-062419


Tot Ion: 75 Resp: 22707
 Ion Ratio Lower Upper
 75 100
 110 8.7 18.1 54.3#
 77 0.0 24.6 36.8#



#38
 Carbon Tetrachloride
 Concen: 6.657 ug/l
 RT: 7.66 min Scan# 1832
 Delta R.T. -0.11 min
 Lab File: VN056484.D
 Acq: 25 Jun 2019 14:18

Tgt Ion: 117 Resp: 31017
 Ion Ratio Lower Upper
 117 100
 119 4.8 79.1 118.7#
 121 0.0 24.8 37.2#





#43

Isopropyl Acetate

Concen: 15.041 ug/l

RT: 8.28 min Scan# 2025

Delta R.T. 0.12 min

Lab File: VN056484.D

Acq: 25 Jun 2019 14:18

Instrument :

MSVOA_N

ClientSampled :

TR-04-062419


Tot Ion: 43 Resp: 121435

Ion Ratio Lower Upper

43 100

61 0.0 22.0 33.0#




87 0.0 11.0 16.4#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 61.00 (60.70 to 61.70): VN0

Ion 87.00 (86.70 to 87.70): VN0



#44

Trichloroethene

Concen: 0.544 ug/l

RT: 8.84 min Scan# 2197

Delta R.T. 0.00 min

Lab File: VN056484.D


Acq: 25 Jun 2019 14:18

Tgt Ion:130 Resp: 2403

Ion Ratio Lower Upper



130 100


95 77.1 0.0 191.0



Abundance Ion 129.80 (129.50 to 130.50): V

Ion 94.90 (94.60 to 95.60): VN0









#45
 1,2-Dichloropropane
 Concen: 0.430 ug/l
 RT: 9.27 min Scan# 2331
 Delta R.T. 0.15 min
 Lab File: VN056484.D
 Acq: 25 Jun 2019 14:18




Instrument :
 MSVOA_N
 ClientSampleId :
 TR-04-062419


Tot Ion: 63 Resp: 1884
 Ion Ratio Lower Upper
 63 100
 65 0.0 25.0 37.6#



#48
 Methyl methacrylate
 Concen: 2.172 ug/l
 RT: 9.18 min Scan# 2304
 Delta R.T. -0.02 min
 Lab File: VN056484.D
 Acq: 25 Jun 2019 14:18

Tgt Ion: 41 Resp: 8789
 Ion Ratio Lower Upper
 41 100
 69 0.0 72.6 109.0#
 39 0.0 42.6 63.8#





#49

1,4-Dioxane

Concen: 0.357 ug/l

RT: 9.26 min Scan# 2328

Delta R.T. 0.06 min

Lab File: VN056484.D

Acq: 25 Jun 2019 14:18

Instrument :

MSVOA_N

ClientSampleId :

TR-04-062419


Tot Ion: 88 Resp: 30

Ion Ratio Lower Upper

88 100

43 4585843.3 22.0 33.0#




58 19543.3 51.8 77.8#



Abundance Ion 88.00 (87.70 to 88.70): VN0

Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0



#50

Toluene-d8

Concen: 51.582 ug/l

RT: 10.09 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN056484.D


Acq: 25 Jun 2019 14:18

Tgt Ion: 98 Resp: 757896

Ion Ratio Lower Upper



98 100


100 64.1 52.2 78.2



Abundance Ion 98.00 (97.70 to 98.70): VN0

Ion 100.00 (99.70 to 100.70): VN0





#51

4-Methyl-2-Pentanone

Concen: 1.520 ug/l

RT: 10.02 min Scan# 2566

Delta R.T. 0.04 min

Lab File: VN056484.D

Acq: 25 Jun 2019 14:18

Instrument :

MSVOA_N

ClientSampleId :


TR-04-062419

Tot Ion: 43 Resp: 7921

Ion Ratio Lower Upper

43 100

58 0.0 31.8 47.8#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0

6000

5000

4000

3000

2000



1000

0

10.00

10.05

Time-->



#54

cis-1,3-Dichloropropene

Concen: 10.483 ug/l

RT: 9.90 min Scan# 2529

Delta R.T. 0.07 min

Lab File: VN056484.D

Acq: 25 Jun 2019 14:18


Tgt Ion: 75 Resp: 61393

Ion Ratio Lower Upper

75 100

77 0.6 25.2 37.8#

39 27.1 34.2 51.4#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 76.90 (76.60 to 77.60): VN0

Ion 39.00 (38.70 to 39.70): VN0

40000

30000

20000

10000

0


9.85


9.90

9.95

10.00

Time-->









#57
 1,3-Dichloropropane
 Concen: 1.794 ug/l
 RT: 10.76 min Scan# 2796
 Delta R.T. 0.06 min
 Lab File: VN056484.D
 Acq: 25 Jun 2019 14:18




Instrument :
 MSVOA_N
 ClientSampled :
 TR-04-062419


Tot Ion: 76 Resp: 12136
 Ion Ratio Lower Upper
 76 100
 78 0.0 25.9 38.9#



#59
 2-Hexanone
 Concen: 41.870 ug/l
 RT: 10.76 min Scan# 2796
 Delta R.T. 0.01 min
 Lab File: VN056484.D
 Acq: 25 Jun 2019 14:18

Tgt Ion: 43 Resp: 160354
 Ion Ratio Lower Upper
 43 100
 58 19.9 27.8 83.4#









#62
4-Bromofluorobenzene
Concen: 49.381 ug/l
RT: 12.40 min Scan# 3305
Delta R.T. 0.00 min
Lab File: VN056484.D
Acq: 25 Jun 2019 14:18




Instrument :
MSVOA_N
ClientSampleId :
TR-04-062419


Tot Ion: 95 Resp: 241871
Ion Ratio Lower Upper
95 100
174 90.3 0.0 179.4
176 87.0 0.0 173.4



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. 0.00 min
Lab File: VN056484.D
Acq: 25 Jun 2019 14:18

Tgt Ion: 117 Resp: 603732
Ion Ratio Lower Upper
117 100
82 52.9 42.0 63.0
119 32.0 25.3 37.9





#71

Bromoform

Concen: 3.599 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. 0.27 min

Lab File: VN056484.D

Acq: 25 Jun 2019 14:18

Instrument :

MSVOA_N

ClientSampled :

TR-04-062419


Tot Ion:173 Resp: 1612

Ion Ratio Lower Upper

173 100

175 880.3 24.4 73.4#




254 0.0 0.0 0.0



Abundance Ion 172.70 (172.40 to 173.40): V

Ion 174.70 (174.40 to 175.40): V

Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.34 min Scan# 3598

Delta R.T. 0.00 min

Lab File: VN056484.D

Acq: 25 Jun 2019 14:18


Tgt Ion:152 Resp: 169019

Ion Ratio Lower Upper

152 100

115 56.1 28.1 84.4



150 157.8 0.0 348.4




Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V





#73

Isopropylbenzene

Concen: Below Cal

RT: 12.40 min Scan# 3305

Delta R.T. 0.15 min

Lab File: VN056484.D

Acq: 25 Jun 2019 14:18

Instrument :

MSVOA_N

ClientSampled :





TR-04-062419

Tot Ion:105 Resp: 148

Ion Ratio Lower Upper

105 100

120 0.0 13.4 40.2#



#74

N-amyl acetate

Concen: 2.414 ug/l

RT: 11.88 min Scan# 3142

Delta R.T. -0.19 min

Lab File: VN056484.D

Acq: 25 Jun 2019 14:18

Tgt Ion: 43 Resp: 12983




Ion Ratio Lower Upper


43 100

70 0.2 34.9 52.3#

55 3.2 20.4 30.6#

61 11.2 19.8 29.8#





#76

1,2,3-Trichloropropane

Concen: 31.650 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. -0.15 min

Lab File: VN056484.D

Acq: 25 Jun 2019 14:18

Instrument :

MSVOA_N

ClientSampled :


TR-04-062419

Tot Ion: 75 Resp: 110921

Ion Ratio Lower Upper




75 100

77 1.2 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 77.00 (76.70 to 77.70): VN0



#79

2-Chlorotoluene

Concen: Below Cal

RT: 12.73 min Scan# 3409

Delta R.T. 0.06 min

Lab File: VN056484.D


Acq: 25 Jun 2019 14:18

Tgt Ion: 91 Resp: 272

Ion Ratio Lower Upper



91 100


126 0.0 17.6 52.8#



Abundance Ion 91.00 (90.70 to 91.70): VN0

Ion 125.90 (125.60 to 126.60): V









#80
 1,3,5-Trimethylbenzene
 Concen: Below Cal
 RT: 12.73 min Scan# 3408
 Delta R.T. 0.00 min
 Lab File: VN056484.D
 Acq: 25 Jun 2019 14:18




Instrument :
 MSVOA_N
 ClientSampled :
 TR-04-062419


Tot Ion:105 Resp: 3814
 Ion Ratio Lower Upper
 105 100
 120 45.2 24.8 74.4



#81
 trans-1,4-Dichloro-2-butene
 Concen: 112.488 ug/l
 RT: 12.40 min Scan# 3305
 Delta R.T. 0.10 min
 Lab File: VN056484.D
 Acq: 25 Jun 2019 14:18

Tgt Ion: 75 Resp: 110921
 Ion Ratio Lower Upper
 75 100
 53 0.0 83.9 125.9#
 89 0.0 35.0 52.6#









#83
 tert-Butylbenzene
 Concen: Below Cal
 RT: 12.73 min Scan# 3408
 Delta R.T. -0.26 min
 Lab File: VN056484.D
 Acq: 25 Jun 2019 14:18

Instrument :
 MSVOA_N
 ClientSampled :
 TR-04-062419

Tot Ion:119 Resp: 415
 Ion Ratio Lower Upper
 119 100
 91 65.5 31.6 94.7
 134 0.0 13.2 39.5#



#95
 Naphthalene
 Concen: 5.277 ug/l
 RT: 15.13 min Scan# 4155
 Delta R.T. 0.00 min
 Lab File: VN056484.D
 Acq: 25 Jun 2019 14:18

Tgt Ion:128 Resp: 280
 Ion Ratio Lower Upper
 128 100
 127 0.0 10.2 15.2#
 129 0.0 8.6 12.8#

