

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062519\  
 Data File : VN056477.D  
 Acq On : 25 Jun 2019 11:34  
 Operator : JC/SP  
 Sample : PB120881ZHE#09  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 PB120881ZHE#09

Quant Time: Jun 26 04:15:21 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N061919W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Jun 21 03:45:40 2019  
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	355844	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	597297	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	606181	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	183262	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	202406	51.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.64%	
35) Dibromofluoromethane	7.59	113	178313	47.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.70%	
50) Toluene-d8	10.09	98	753354	51.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.92%	
62) 4-Bromofluorobenzene	12.40	95	249026	51.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.04%	
Target Compounds						
43) Isopropyl Acetate	8.17	43	17213	2.161	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062519\  
 Data File : VN056477.D  
 Acq On : 25 Jun 2019 11:34  
 Operator : JC/SP  
 Sample : PB120881ZHE#09  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 PB120881ZHE#09

Quant Time: Jun 26 04:15:21 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N061919W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Jun 21 03:45:40 2019  
 Response via : Initial Calibration

