

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062519\
 Data File : VN056487.D
 Acq On : 25 Jun 2019 15:28
 Operator : JC/SP
 Sample : K3512-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SAMPLE-7

Quant Time: Jun 26 04:53:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 03:45:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	352203	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	607313	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	615922	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	176006	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	204958	52.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.02%	
35) Dibromofluoromethane	7.59	113	177664	46.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.78%	
50) Toluene-d8	10.09	98	771398	52.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.66%	
62) 4-Bromofluorobenzene	12.40	95	248270	50.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.04%	
Target Compounds						
16) Acetone	3.82	43	22811	11.664	ug/l	96
18) Methyl Acetate	4.33	43	10600	1.933	ug/l #	86
20) Methylene Chloride	4.55	84	7953	1.290	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062519\
Data File : VN056487.D
Acq On : 25 Jun 2019 15:28
Operator : JC/SP
Sample : K3512-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SAMPLE-7

Quant Time: Jun 26 04:53:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 03:45:40 2019
Response via : Initial Calibration

