


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN062519\
Data File : VN056469.D
Acq On : 25 Jun 2019 8:09
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
Title : SW846 8260
Last Update : Fri Jun 21 03:45:40 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.7	8726	PASS
75	95	30	60	46.3	24242	PASS
95	95	100	100	100.0	52309	PASS
96	95	5	9	6.7	3494	PASS
173	174	0.00	2	1.1	517	PASS
174	95	50	100	91.0	47576	PASS
175	174	5	9	7.0	3329	PASS
176	174	95	101	96.7	46029	PASS
177	176	5	9	6.7	3075	PASS