

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062520\
 Data File : VN062122.D
 Acq On : 25 Jun 2020 19:42
 Operator : JC/MD
 Sample : L3087-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FM1296

Quant Time: Jun 26 04:11:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	241340	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	432652	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	410692	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	156034	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	154592	45.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.30%	
35) Dibromofluoromethane	7.55	113	135600	48.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.70%	
50) Toluene-d8	10.06	98	529867	47.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.52%	
62) 4-Bromofluorobenzene	12.38	95	178672	47.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.32%	
Target Compounds						
16) Acetone	3.78	43	23329	18.189	ug/l	99
20) Methylene Chloride	4.51	84	6087	1.749	ug/l	90
30) Chloroform	7.33	83	16815	2.988	ug/l	98
95) Naphthalene	15.10	128	5568	4.684	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062520\
Data File : VN062122.D
Acq On : 25 Jun 2020 19:42
Operator : JC/MD
Sample : L3087-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FM1296

Quant Time: Jun 26 04:11:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

