

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062520\
 Data File : VN062124.D
 Acq On : 25 Jun 2020 20:30
 Operator : JC/MD
 Sample : L3043-18
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE120D1-20200615

Quant Time: Jun 26 04:16:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	241525	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	440150	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	426275	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	159362	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	157146	46.36	ug/l	0.00
Spiked Amount			Recovery	=	92.72%	
35) Dibromofluoromethane	7.55	113	139154	49.28	ug/l	0.00
Spiked Amount			Recovery	=	98.56%	
50) Toluene-d8	10.06	98	547122	48.47	ug/l	0.00
Spiked Amount			Recovery	=	96.94%	
62) 4-Bromofluorobenzene	12.38	95	185857	48.74	ug/l	0.00
Spiked Amount			Recovery	=	97.48%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	3.72	101	49188	17.731	ug/l	98
12) 1,1-Dichloroethene	3.70	96	23725	8.555	ug/l	89
24) 1,1-Dichloroethane	5.80	63	7318	1.322	ug/l #	88
27) cis-1,2-Dichloroethene	6.78	96	10578	2.975	ug/l	90
30) Chloroform	7.33	83	3250	0.577	ug/l	94
38) Carbon Tetrachloride	7.73	117	1812	0.434	ug/l #	92
44) Trichloroethene	8.80	130	2867469	842.988	ug/l	99
49) 1,4-Dioxane	9.17	88	1564	26.338	ug/l	92
55) 1,1,2-Trichloroethane	10.54	97	2459	0.760	ug/l	91
64) Tetrachloroethene	10.60	164	13538	4.649	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062520\
Data File : VN062124.D
Acq On : 25 Jun 2020 20:30
Operator : JC/MD
Sample : L3043-18
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE120D1-20200615

Quant Time: Jun 26 04:16:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

