

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN062520\
 Data File : VN062125.D
 Acq On : 25 Jun 2020 20:54
 Operator : JC/MD
 Sample : L3043-19
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE120D2-20200615

Quant Time: Jun 30 14:42:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	244585	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	447420	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	431989	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	161005	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	158029	46.04	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	92.08%	
35) Dibromofluoromethane	7.55	113	138716	48.33	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	96.66%	
50) Toluene-d8	10.06	98	559466	48.76	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	97.52%	
62) 4-Bromofluorobenzene	12.38	95	184191	47.51	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	95.02%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	3.72	101	65766	23.410	ug/l	100
12) 1,1-Dichloroethene	3.70	96	4995	1.779	ug/l	91
27) cis-1,2-Dichloroethene	6.79	96	10884	3.023	ug/l	87
30) Chloroform	7.33	83	3098	0.543	ug/l	86
38) Carbon Tetrachloride	7.73	117	2911	0.687	ug/l	97
44) Trichloroethene	8.80	130	2335617	675.476	ug/l	99
64) Tetrachloroethene	10.60	164	19691	6.673	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN062520\
Data File : VN062125.D
Acq On : 25 Jun 2020 20:54
Operator : JC/MD
Sample : L3043-19
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE120D2-20200615

Quant Time: Jun 30 14:42:17 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

