


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN062520\
 Data File : VN062099.D
 Acq On : 25 Jun 2020 9:44
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Title : SW846 8260
 Last Update : Wed Jun 24 15:33:00 2020



AutoFind: Scans 3296, 3297, 3298; Background Corrected with Scan 3283

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.0	12098	PASS
75	95	30	60	49.4	31362	PASS
95	95	100	100	100.0	63528	PASS
96	95	5	9	7.3	4612	PASS
173	174	0.00	2	1.6	804	PASS
174	95	50	100	76.8	48765	PASS
175	174	5	9	7.9	3865	PASS
176	174	95	101	95.7	46680	PASS
177	176	5	9	6.5	3052	PASS