

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062521\
 Data File : VN067520.D
 Acq On : 25 Jun 2021 21:38
 Operator : JC/MD
 Sample : M2806-07MS 10X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Jun 26 00:41:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062421W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 24 12:18:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	192867	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	453522	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	418194	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	160364	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	235203	59.59	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	119.18%
35) Dibromofluoromethane	8.021	113	143740	51.45	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.90%
50) Toluene-d8	10.443	98	623608	52.33	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.66%
62) 4-Bromofluorobenzene	12.734	95	225359	53.08	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	106.16%

Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.070	85	190441	58.79	ug/l	99
3) Chloromethane	2.298	50	226927	62.68	ug/l	99
4) Vinyl Chloride	2.440	62	210115	65.06	ug/l	97
5) Bromomethane	2.837	94	107169	60.09	ug/l	99
6) Chloroethane	3.009	64	138336	66.39	ug/l	97
7) Trichlorofluoromethane	3.368	101	302318	61.87	ug/l	98
8) Diethyl Ether	3.819	74	121081	60.38	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.202	101	151223	56.61	ug/l	92
10) Methyl Iodide	4.411	142	147208	50.05	ug/l	97
11) Tert butyl alcohol	5.388	59	174667	270.95	ug/l	99
12) 1,1-Dichloroethene	4.173	96	155647	58.92	ug/l	97
13) Acrolein	4.039	56	159020	290.15	ug/l	99
14) Allyl chloride	4.832	41	329313	56.84	ug/l	95
15) Acrylonitrile	5.557	53	535498	312.00	ug/l	100
16) Acetone	4.288	43	443543	293.59	ug/l	90
17) Carbon Disulfide	4.521	76	485206	56.36	ug/l	99
18) Methyl Acetate	4.857	43	244255	57.93	ug/l	100
19) Methyl tert-butyl Ether	5.621	73	635566	60.68	ug/l	94
20) Methylene Chloride	5.090	84	199278	56.06	ug/l	97
21) trans-1,2-Dichloroethene	5.597	96	162586	57.28	ug/l	92
22) Diisopropyl ether	6.501	45	716839	62.92	ug/l	96
23) Vinyl Acetate	6.436	43	3059676	316.54	ug/l	99
24) 1,1-Dichloroethane	6.385	63	387856	62.38	ug/l	98
25) 2-Butanone	7.340	43	731890	311.61	ug/l	100
26) 2,2-Dichloropropane	7.329	77	219423	41.46	ug/l	97
27) cis-1,2-Dichloroethene	7.327	96	194345	57.95	ug/l	98
28) Bromochloromethane	7.659	49	181798	61.35	ug/l #	82
29) Tetrahydrofuran	7.689	42	488102	328.19	ug/l	99
30) Chloroform	7.820	83	356822	60.25	ug/l	100
31) Cyclohexane	8.094	56	360159	57.01	ug/l	97
32) 1,1,1-Trichloroethane	8.016	97	292008	58.93	ug/l	96
36) 1,1-Dichloropropene	8.223	75	278483	54.51	ug/l	97
37) Ethyl Acetate	7.418	43	297285	55.09	ug/l	98
38) Carbon Tetrachloride	8.206	117	232665	53.47	ug/l	99
39) Methylcyclohexane	9.462	83	302912	51.44	ug/l	99
40) Benzene	8.461	78	826024	55.03	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062521\
 Data File : VN067520.D
 Acq On : 25 Jun 2021 21:38
 Operator : JC/MD
 Sample : M2806-07MS 10X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Jun 26 00:41:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062421W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 24 12:18:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.638	41	152327	55.81	ug/l	96
42) 1,2-Dichloroethane	8.531	62	304329	55.74	ug/l	99
43) Isopropyl Acetate	8.563	43	533392	57.31	ug/l	99
44) Trichloroethene	9.218	130	152037	51.29	ug/l	89
45) 1,2-Dichloropropane	9.491	63	232781	56.61	ug/l	98
46) Dibromomethane	9.580	93	135006	54.18	ug/l #	90
47) Bromodichloromethane	9.762	83	277231	54.32	ug/l	99
48) Methyl methacrylate	9.561	41	230555	57.21	ug/l	97
49) 1,4-Dioxane	9.572	88	63821	1110.04	ug/l	99
51) 4-Methyl-2-Pentanone	10.333	43	1602129	294.82	ug/l	100
52) Toluene	10.508	92	468966	53.50	ug/l	96
53) t-1,3-Dichloropropene	10.719	75	305830	47.74	ug/l	97
54) cis-1,3-Dichloropropene	10.191	75	332429	52.75	ug/l	93
55) 1,1,2-Trichloroethane	10.899	97	183859	54.73	ug/l	96
56) Ethyl methacrylate	10.765	69	342444	56.57	ug/l	98
57) 1,3-Dichloropropane	11.047	76	359795	56.61	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.065	63	175	17.33	ug/l #	50
59) 2-Hexanone	11.087	43	1133347	296.14	ug/l	98
60) Dibromochloromethane	11.240	129	163486	47.65	ug/l	98
61) 1,2-Dibromoethane	11.350	107	177631	54.36	ug/l	100
64) Tetrachloroethene	10.980	164	120389	44.26	ug/l	93
65) Chlorobenzene	11.773	112	450293	51.81	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.843	131	149434	47.83	ug/l	97
67) Ethyl Benzene	11.846	91	952315	53.80	ug/l	98
68) m/p-Xylenes	11.953	106	703299	117.35	ug/l	100
69) o-Xylene	12.283	106	323820	54.54	ug/l	99
70) Styrene	12.296	104	520973	53.62	ug/l	99
71) Bromoform	12.460	173	94472	42.22	ug/l #	100
73) Isopropylbenzene	12.581	105	1080181	66.17	ug/l	98
74) N-amyl acetate	12.388	43	455670	57.82	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.830	83	288555	54.65	ug/l	99
76) 1,2,3-Trichloropropane	12.884	75	375834	79.63	ug/l #	100
77) Bromobenzene	12.862	156	144142	47.90	ug/l	76
78) n-propylbenzene	12.921	91	1600874	79.52	ug/l	96
79) 2-Chlorotoluene	13.010	91	664677	54.99	ug/l	95
80) 1,3,5-Trimethylbenzene	13.061	105	962150	72.17	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.629	75	81401	47.19	ug/l	95
82) 4-Chlorotoluene	13.106	91	619901	52.24	ug/l	96
83) tert-Butylbenzene	13.326	119	587378	55.80	ug/l	99
84) 1,2,4-Trimethylbenzene	13.372	105	1845867	141.13	ug/l	98
85) sec-Butylbenzene	13.503	105	972170	61.69	ug/l	98
86) p-Isopropyltoluene	13.619	119	625723	53.38	ug/l	98
87) 1,3-Dichlorobenzene	13.616	146	270200	49.50	ug/l	97
88) 1,4-Dichlorobenzene	13.696	146	266630	48.84	ug/l	97
89) n-Butylbenzene	13.943	91	707269	60.23	ug/l	98
90) Hexachloroethane	14.214	117	131757	58.98	ug/l	76
91) 1,2-Dichlorobenzene	13.989	146	273217	51.25	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	14.603	75	56130	59.66	ug/l	90
93) 1,2,4-Trichlorobenzene	15.265	180	117992	44.07	ug/l	97
94) Hexachlorobutadiene	15.373	225	52266	43.29	ug/l	98
95) Naphthalene	15.507	128	417069	46.83	ug/l	98
96) 1,2,3-Trichlorobenzene	15.703	180	115267	45.16	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062521\
Data File : VN067520.D
Acq On : 25 Jun 2021 21:38
Operator : JC/MD
Sample : M2806-07MS 10X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :


Quant Time: Jun 26 00:41:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062421W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 24 12:18:45 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 26 00:41:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062421W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 24 12:18:45 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 8.088 min Scan# 2274
Delta R.T. 0.002 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:168 Resp: 192867
Ion Ratio Lower Upper
168 100
99 90.8 62.9 94.3



#2
Dichlorodifluoromethane
Concen: 58.79 ug/l
RT: 2.070 min Scan# 30
Delta R.T. 0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion: 85 Resp: 190441
Ion Ratio Lower Upper
85 100
87 31.4 16.1 48.2





#3

Chloromethane

Concen: 62.68 ug/l

RT: 2.298 min Scan# 115

Delta R.T. -0.003 min

Lab File: VN067520.D





Acq: 25 Jun 2021 21:38

Tgt Ion: 50 Resp: 226927

Ion Ratio Lower Upper

50 100

52 32.3 25.6 38.4



#4

Vinyl Chloride

Concen: 65.06 ug/l

RT: 2.440 min Scan# 168

Delta R.T. -0.003 min

Lab File: VN067520.D




Acq: 25 Jun 2021 21:38


Tgt Ion: 62 Resp: 210115

Ion Ratio Lower Upper

62 100

64 32.0 24.4 36.6









#5
Bromomethane
Concen: 60.09 ug/l
RT: 2.837 min Scan# 316
Delta R.T. 0.005 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 94 Resp: 107169
Ion Ratio Lower Upper
94 100
96 93.1 73.7 110.5



#6
Chloroethane
Concen: 66.39 ug/l
RT: 3.009 min Scan# 380
Delta R.T. 0.003 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion: 64 Resp: 138336
Ion Ratio Lower Upper
64 100
66 32.7 24.7 37.1









#7
 Trichlorofluoromethane
 Concen: 61.87 ug/l
 RT: 3.368 min Scan# 514
 Delta R.T. 0.000 min
 Lab File: VN067520.D
 Acq: 25 Jun 2021 21:38




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:101 Resp: 302318
 Ion Ratio Lower Upper
 101 100
 103 65.6 51.2 76.8



#8
 Diethyl Ether
 Concen: 60.38 ug/l
 RT: 3.819 min Scan# 682
 Delta R.T. -0.002 min
 Lab File: VN067520.D
 Acq: 25 Jun 2021 21:38

Tgt Ion: 74 Resp: 121081
 Ion Ratio Lower Upper
 74 100
 45 102.4 50.2 150.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 56.61 ug/l
RT: 4.202 min Scan# 825
Delta R.T. 0.003 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:101 Resp: 151223
Ion Ratio Lower Upper
101 100
85 50.7 39.6 59.4
151 57.7 54.1 81.1



#10
Methyl Iodide
Concen: 50.05 ug/l
RT: 4.411 min Scan# 903
Delta R.T. -0.003 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion:142 Resp: 147208
Ion Ratio Lower Upper
142 100
127 51.5 43.1 64.7
141 14.6 11.4 17.0









#11
Tert butyl alcohol
Concen: 270.95 ug/l
RT: 5.388 min Scan# 1267
Delta R.T. -0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 59 Resp: 174667
Ion Ratio Lower Upper
59 100
57 10.4 8.7 13.1



#12
1,1-Dichloroethene
Concen: 58.92 ug/l
RT: 4.173 min Scan# 814
Delta R.T. -0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion: 96 Resp: 155647
Ion Ratio Lower Upper
96 100
61 195.6 152.6 229.0
98 66.2 50.9 76.3









#13
Acrolein
Concen: 290.15 ug/l
RT: 4.039 min Scan# 764
Delta R.T. -0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 56 Resp: 159020
Ion Ratio Lower Upper
56 100
55 71.0 57.4 86.0



#14
Allyl chloride
Concen: 56.84 ug/l
RT: 4.832 min Scan# 1060
Delta R.T. -0.003 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion: 41 Resp: 329313
Ion Ratio Lower Upper
41 100
39 71.2 60.9 91.3
76 31.3 24.3 36.5





#15

Acrylonitrile

Concen: 312.00 ug/l

RT: 5.557 min Scan# 1330





Delta R.T. 0.001 min

Lab File: VN067520.D

Acq: 25 Jun 2021 21:38

Tgt Ion:	53	Resp:	535498
Ion	Ratio	Lower	Upper
53	100		
52	82.5	66.2	99.2
51	37.6	30.2	45.4

Instrument :
MSVOA_N
ClientSampleId :



#16

Acetone

Concen: 293.59 ug/l




RT: 4.288 min Scan# 857


Delta R.T. -0.003 min

Lab File: VN067520.D

Acq: 25 Jun 2021 21:38

Tgt Ion:	43	Resp:	443543
Ion	Ratio	Lower	Upper
43	100		
58	31.4	21.2	31.8









#17
Carbon Disulfide
Concen: 56.36 ug/l
RT: 4.521 min Scan# 944
Delta R.T. 0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 76 Resp: 485206
Ion Ratio Lower Upper
76 100
78 9.1 7.7 11.5



#18
Methyl Acetate
Concen: 57.93 ug/l
RT: 4.857 min Scan# 1069
Delta R.T. -0.002 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion: 43 Resp: 244255
Ion Ratio Lower Upper
43 100
74 21.9 17.4 26.0









#19

Methyl tert-butyl Ether
Concen: 60.68 ug/l
RT: 5.621 min Scan# 1354
Delta R.T. 0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Instrument :
MSVOA_N
ClientSampled :




Tgt Ion: 73 Resp: 635566
Ion Ratio Lower Upper
73 100
57 24.2 17.0 25.6




#20

Methylene Chloride
Concen: 56.06 ug/l
RT: 5.090 min Scan# 1156
Delta R.T. -0.003 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion: 84 Resp: 199278
Ion Ratio Lower Upper
84 100
49 143.0 110.9 166.3
51 43.2 32.6 49.0
86 62.8 51.6 77.4









#21
trans-1,2-Dichloroethene
Concen: 57.28 ug/l
RT: 5.597 min Scan# 1345
Delta R.T. 0.003 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 96 Resp: 162586
Ion Ratio Lower Upper
96 100
61 168.6 123.5 185.3
98 61.9 48.6 72.8



#22
Diisopropyl ether
Concen: 62.92 ug/l
RT: 6.501 min Scan# 1682
Delta R.T. -0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion: 45 Resp: 716839
Ion Ratio Lower Upper
45 100
43 58.1 49.4 74.2
87 23.8 20.9 31.3
59 11.6 9.0 13.4









#23
 Vinyl Acetate
 Concen: 316.54 ug/l
 RT: 6.436 min Scan# 1658
 Delta R.T. 0.000 min
 Lab File: VN067520.D
 Acq: 25 Jun 2021 21:38




Instrument :
 MSVOA_N
 ClientSampled :


Tgt Ion: 43 Resp: 3059676
 Ion Ratio Lower Upper
 43 100
 86 8.6 7.3 10.9



#24
 1,1-Dichloroethane
 Concen: 62.38 ug/l
 RT: 6.385 min Scan# 1639
 Delta R.T. -0.003 min
 Lab File: VN067520.D
 Acq: 25 Jun 2021 21:38

Tgt Ion: 63 Resp: 387856
 Ion Ratio Lower Upper
 63 100
 98 5.1 3.0 9.0
 100 3.7 1.9 5.7









#25
2-Butanone
Concen: 311.61 ug/l
RT: 7.340 min Scan# 1995
Delta R.T. 0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 43 Resp: 731890
Ion Ratio Lower Upper
43 100
72 24.7 19.7 29.5



#26
2,2-Dichloropropane
Concen: 41.46 ug/l
RT: 7.329 min Scan# 1991
Delta R.T. 0.005 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion: 77 Resp: 219423
Ion Ratio Lower Upper
77 100
97 20.9 9.8 29.3









#27
 cis-1,2-Dichloroethene
 Concen: 57.95 ug/l
 RT: 7.327 min Scan# 1990
 Delta R.T. 0.003 min
 Lab File: VN067520.D
 Acq: 25 Jun 2021 21:38




Instrument :
 MSVOA_N
 ClientSampled :


Tgt Ion:	96	Resp:	194345
Ion	Ratio	Lower	Upper
96	100		
61	171.5	0.0	338.8
98	63.0	0.0	129.8



#28
 Bromochloromethane
 Concen: 61.35 ug/l
 RT: 7.659 min Scan# 2114
 Delta R.T. 0.003 min
 Lab File: VN067520.D
 Acq: 25 Jun 2021 21:38

Tgt Ion:	49	Resp:	181798
Ion	Ratio	Lower	Upper
49	100		
129	1.5	0.0	3.8
130	47.5	49.7	74.5#





#29

Tetrahydrofuran

Concen: 328.19 ug/l

RT: 7.689 min Scan# 2125





Delta R.T. -0.000 min

Lab File: VN067520.D

Acq: 25 Jun 2021 21:38

Instrument :
MSVOA_N
ClientSampleId :

Tgt Ion:	42	Resp:	488102
Ion Ratio	Lower	Upper	
42	100		
72	40.0	32.8	49.2
71	38.2	31.2	46.8



#30

Chloroform

Concen: 60.25 ug/l




RT: 7.820 min Scan# 2174


Delta R.T. 0.003 min

Lab File: VN067520.D

Acq: 25 Jun 2021 21:38

Tgt Ion:	83	Resp:	356822
Ion Ratio	Lower	Upper	
83	100		
85	64.7	51.5	77.3









#31
Cyclohexane
Concen: 57.01 ug/l
RT: 8.094 min Scan# 2276
Delta R.T. -0.002 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 56 Resp: 360159
Ion Ratio Lower Upper
56 100
69 31.4 27.3 40.9
84 78.5 64.9 97.3



#32
1,1,1-Trichloroethane
Concen: 58.93 ug/l
RT: 8.016 min Scan# 2247
Delta R.T. 0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion: 97 Resp: 292008
Ion Ratio Lower Upper
97 100
99 63.4 50.2 75.2
61 52.7 38.2 57.2









#33
 1,2-Dichloroethane-d4
 Concen: 59.59 ug/l
 RT: 8.440 min Scan# 2405
 Delta R.T. -0.000 min
 Lab File: VN067520.D
 Acq: 25 Jun 2021 21:38

Instrument :
 MSVOA_N
 ClientSampled :

Tgt Ion: 65 Resp: 235203




Ion	Ratio	Lower	Upper
65	100		
67	53.2	0.0	98.6




#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 8.968 min Scan# 2602
 Delta R.T. -0.003 min
 Lab File: VN067520.D
 Acq: 25 Jun 2021 21:38

Tgt Ion: 114 Resp: 453522

Ion	Ratio	Lower	Upper
114	100		
63	28.3	0.0	49.0
88	18.7	0.0	36.6









#35
Dibromofluoromethane
Concen: 51.45 ug/l
RT: 8.021 min Scan# 2249
Delta R.T. -0.003 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:113 Resp: 143740
Ion Ratio Lower Upper
113 100
111 106.3 83.8 125.8
192 12.0 13.4 20.0#



#36
1,1-Dichloropropene
Concen: 54.51 ug/l
RT: 8.223 min Scan# 2324
Delta R.T. 0.001 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion: 75 Resp: 278483
Ion Ratio Lower Upper
75 100
110 28.4 15.7 47.0
77 30.1 24.6 37.0





#37

Ethyl Acetate

Concen: 55.09 ug/l

RT: 7.418 min Scan# 2024





Delta R.T. -0.000 min

Lab File: VN067520.D

Acq: 25 Jun 2021 21:38

Tgt Ion:	43	Resp:	297285
Ion	Ratio	Lower	Upper
43	100		
61	14.0	10.6	16.0
70	9.6	8.1	12.1

Instrument :
MSVOA_N
ClientSampled :



#38

Carbon Tetrachloride

Concen: 53.47 ug/l




RT: 8.206 min Scan# 2318


Delta R.T. 0.000 min

Lab File: VN067520.D

Acq: 25 Jun 2021 21:38

Tgt Ion:	117	Resp:	232665
Ion	Ratio	Lower	Upper
117	100		
119	92.6	75.4	113.2
121	29.5	23.5	35.3





#39

Methylcyclohexane

Concen: 51.44 ug/l

RT: 9.462 min Scan# 2786

Delta R.T. 0.001 min





Lab File: VN067520.D

Acq: 25 Jun 2021 21:38

Tgt Ion: 83 Resp: 302912

Ion	Ratio	Lower	Upper
83	100		
55	83.5	66.6	99.8
98	42.9	35.3	52.9

Instrument :
MSVOA_N
ClientSampleId :



#40

Benzene

Concen: 55.03 ug/l

RT: 8.461 min Scan# 2413




Delta R.T. 0.000 min


Lab File: VN067520.D

Acq: 25 Jun 2021 21:38

Tgt Ion: 78 Resp: 826024

Ion	Ratio	Lower	Upper
78	100		
77	23.2	18.9	28.3









#41
Methacrylonitrile
Concen: 55.81 ug/l
RT: 7.638 min Scan# 2106
Delta R.T. 0.003 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Instrument :
MSVOA_N
ClientSampled :

Tgt Ion: 41 Resp: 152327




Ion	Ratio	Lower	Upper
41	100		
39	61.3	47.6	71.4
67	77.8	57.7	86.5
52	33.0	26.2	39.4




#42
1,2-Dichloroethane
Concen: 55.74 ug/l
RT: 8.531 min Scan# 2439
Delta R.T. -0.002 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion: 62 Resp: 304329

Ion	Ratio	Lower	Upper
62	100		
98	7.3	0.0	14.2









#43
Isopropyl Acetate
Concen: 57.31 ug/l
RT: 8.563 min Scan# 2451
Delta R.T. 0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 43 Resp: 533392
Ion Ratio Lower Upper
43 100
61 23.7 19.0 28.4
87 11.1 9.5 14.3



#44
Trichloroethene
Concen: 51.29 ug/l
RT: 9.218 min Scan# 2695
Delta R.T. 0.001 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion:130 Resp: 152037
Ion Ratio Lower Upper
130 100
95 128.9 0.0 234.4





#45

1,2-Dichloropropane

Concen: 56.61 ug/l

RT: 9.491 min Scan# 2797

Delta R.T. -0.003 min

Lab File: VN067520.D





Acq: 25 Jun 2021 21:38

Tgt Ion: 63 Resp: 232781

Ion Ratio Lower Upper

63 100

65 31.3 24.3 36.5



#46

Dibromomethane

Concen: 54.18 ug/l

RT: 9.580 min Scan# 2830

Delta R.T. 0.001 min

Lab File: VN067520.D

Acq: 25 Jun 2021 21:38




Tgt Ion: 93 Resp: 135006


Ion Ratio Lower Upper

93 100

95 82.4 66.1 99.1

174 58.4 61.0 91.4#









#47
Bromodichloromethane
Concen: 54.32 ug/l
RT: 9.762 min Scan# 2898
Delta R.T. -0.003 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 83 Resp: 277231
Ion Ratio Lower Upper
83 100
85 62.7 49.9 74.9
127 6.2 6.1 9.1



#48
Methyl methacrylate
Concen: 57.21 ug/l
RT: 9.561 min Scan# 2823
Delta R.T. -0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion: 41 Resp: 230555
Ion Ratio Lower Upper
41 100
69 83.7 64.6 96.8
39 64.1 53.3 79.9





#49

1,4-Dioxane

Concen: 1110.04 ug/l

RT: 9.572 min Scan# 2827

Delta R.T. 0.001 min

Lab File: VN067520.D

Acq: 25 Jun 2021 21:38

Tgt Ion: 88 Resp: 63821

Ion Ratio Lower Upper

88 100





43 37.7 28.8 43.2

58 75.3 60.4 90.6

Instrument :

MSVOA_N

ClientSampleId :



#50

Toluene-d8

Concen: 52.33 ug/l

RT: 10.443 min Scan# 3152

Delta R.T. 0.000 min

Lab File: VN067520.D




Acq: 25 Jun 2021 21:38


Tgt Ion: 98 Resp: 623608

Ion Ratio Lower Upper

98 100

100 61.4 51.7 77.5









#51
4-Methyl-2-Pentanone
Concen: 294.82 ug/l
RT: 10.333 min Scan# 3111
Delta R.T. 0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 43 Resp: 1602129
Ion Ratio Lower Upper
43 100
58 36.1 28.6 43.0



#52
Toluene
Concen: 53.50 ug/l
RT: 10.508 min Scan# 3176
Delta R.T. 0.001 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion: 92 Resp: 468966
Ion Ratio Lower Upper
92 100
91 179.5 139.6 209.4









#53
t-1,3-Dichloropropene
Concen: 47.74 ug/l
RT: 10.719 min Scan# 3255
Delta R.T. -0.003 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 75 Resp: 305830
Ion Ratio Lower Upper
75 100
77 30.0 25.1 37.7



#54
cis-1,3-Dichloropropene
Concen: 52.75 ug/l
RT: 10.191 min Scan# 3058
Delta R.T. 0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion: 75 Resp: 332429
Ion Ratio Lower Upper
75 100
77 30.2 24.6 37.0
39 56.2 50.7 76.1









#55
1,1,2-Trichloroethane
Concen: 54.73 ug/l
RT: 10.899 min Scan# 3322
Delta R.T. -0.003 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:	97	Resp:	183859
Ion	Ratio	Lower	Upper
97	100		
83	95.2	72.0	108.0
85	59.8	46.8	70.2
99	60.3	49.8	74.6



#56
Ethyl methacrylate
Concen: 56.57 ug/l
RT: 10.765 min Scan# 3272
Delta R.T. 0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion:	69	Resp:	342444
Ion	Ratio	Lower	Upper
69	100		
41	72.5	58.6	88.0
39	45.5	38.3	57.5









#57
1,3-Dichloropropane
Concen: 56.61 ug/l
RT: 11.047 min Scan# 3377
Delta R.T. 0.001 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 76 Resp: 359795
Ion Ratio Lower Upper
76 100
78 31.5 25.7 38.5



#58
2-Chloroethyl Vinyl ether
Concen: 17.33 ug/l
RT: 10.065 min Scan# 3011
Delta R.T. 0.022 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion: 63 Resp: 175
Ion Ratio Lower Upper
63 100
106 0.0 19.8 29.8#









#59
2-Hexanone
Concen: 296.14 ug/l
RT: 11.087 min Scan# 3392
Delta R.T. -0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 43 Resp: 1133347
Ion Ratio Lower Upper
43 100
58 50.5 24.7 74.1



#60
Dibromochloromethane
Concen: 47.65 ug/l
RT: 11.240 min Scan# 3449
Delta R.T. -0.002 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion:129 Resp: 163486
Ion Ratio Lower Upper
129 100
127 79.2 38.6 115.8









#61
1,2-Dibromoethane
Concen: 54.36 ug/l
RT: 11.350 min Scan# 3490
Delta R.T. 0.003 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:107 Resp: 177631
Ion Ratio Lower Upper
107 100
109 93.3 75.0 112.4



#62
4-Bromofluorobenzene
Concen: 53.08 ug/l
RT: 12.734 min Scan# 4006
Delta R.T. 0.001 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion: 95 Resp: 225359
Ion Ratio Lower Upper
95 100
174 53.0 0.0 122.0
176 49.8 0.0 117.6









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.747 min Scan# 3638
Delta R.T. 0.001 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:117 Resp: 418194
Ion Ratio Lower Upper
117 100
82 70.5 53.1 79.7
119 32.3 25.5 38.3



#64
Tetrachloroethene
Concen: 44.26 ug/l
RT: 10.980 min Scan# 3352
Delta R.T. 0.001 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38


Tgt Ion:164 Resp: 120389
Ion Ratio Lower Upper
164 100
166 126.0 102.1 153.1
129 120.8 86.9 130.3
131 114.1 83.6 125.4








#65
Chlorobenzene
Concen: 51.81 ug/l
RT: 11.773 min Scan# 3648
Delta R.T. 0.002 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Instrument :
MSVOA_N
ClientSampled :







Tgt Ion:112 Resp: 450293
Ion Ratio Lower Upper
112 100
114 31.5 24.6 37.0



#66
1,1,1,2-Tetrachloroethane
Concen: 47.83 ug/l
RT: 11.843 min Scan# 3674
Delta R.T. 0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion:131 Resp: 149434
Ion Ratio Lower Upper
131 100
133 95.6 47.5 142.5
119 70.3 33.1 99.3









#67
Ethyl Benzene
Concen: 53.80 ug/l
RT: 11.846 min Scan# 3675
Delta R.T. -0.002 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 952315
Ion Ratio Lower Upper
91 100
106 28.0 23.4 35.0



#68
m/p-Xylenes
Concen: 117.35 ug/l
RT: 11.953 min Scan# 3715
Delta R.T. -0.003 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion:106 Resp: 703299
Ion Ratio Lower Upper
106 100
91 226.5 180.6 270.8









#69
o-Xylene
Concen: 54.54 ug/l
RT: 12.283 min Scan# 3838
Delta R.T. 0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:106 Resp: 323820
Ion Ratio Lower Upper
106 100
91 237.8 118.3 354.9



#70
Styrene
Concen: 53.62 ug/l
RT: 12.296 min Scan# 3843
Delta R.T. 0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion:104 Resp: 520973
Ion Ratio Lower Upper
104 100
78 59.9 48.2 72.2
103 56.5 44.5 66.7









#71
Bromoform
Concen: 42.22 ug/l
RT: 12.460 min Scan# 3904
Delta R.T. 0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:173 Resp: 94472
Ion Ratio Lower Upper
173 100
175 48.0 23.9 71.8
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.678 min Scan# 4358
Delta R.T. 0.003 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion:152 Resp: 160364
Ion Ratio Lower Upper
152 100
115 87.2 36.1 108.5
150 175.8 0.0 351.8









#73
Isopropylbenzene
Concen: 66.17 ug/l
RT: 12.581 min Scan# 3949
Delta R.T. -0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:105 Resp: 1080181
Ion Ratio Lower Upper
105 100
120 24.1 12.5 37.5



#74
N-ethyl acetate
Concen: 57.82 ug/l
RT: 12.388 min Scan# 3877
Delta R.T. -0.002 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion: 43 Resp: 455670
Ion Ratio Lower Upper
43 100
70 37.8 29.4 44.2
55 21.9 17.0 25.4
61 22.3 17.6 26.4









#75
1,1,2,2-Tetrachloroethane
Concen: 54.65 ug/l
RT: 12.830 min Scan# 4042
Delta R.T. 0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 288555
Ion Ratio Lower Upper
83 100
131 7.4 4.8 14.4
85 63.9 32.2 96.6



#76
1,2,3-Trichloropropane
Concen: 79.63 ug/l
RT: 12.884 min Scan# 4062
Delta R.T. -0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion: 75 Resp: 375834
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0









#77
Bromobenzene
Concen: 47.90 ug/l
RT: 12.862 min Scan# 4054
Delta R.T. 0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:156 Resp: 144142
Ion Ratio Lower Upper
156 100
77 332.7 136.2 408.6
158 98.0 48.4 145.3



#78
n-propylbenzene
Concen: 79.52 ug/l
RT: 12.921 min Scan# 4076
Delta R.T. 0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion: 91 Resp: 1600874
Ion Ratio Lower Upper
91 100
120 18.9 10.4 31.4









#79
2-Chlorotoluene
Concen: 54.99 ug/l
RT: 13.010 min Scan# 4109
Delta R.T. -0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 91 Resp: 664677
Ion Ratio Lower Upper
91 100
126 26.5 14.5 43.5



#80
1,3,5-Trimethylbenzene
Concen: 72.17 ug/l
RT: 13.061 min Scan# 4128
Delta R.T. -0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion: 105 Resp: 962150
Ion Ratio Lower Upper
105 100
120 43.3 22.2 66.6









#81
trans-1,4-Dichloro-2-butene
Concen: 47.19 ug/l
RT: 12.629 min Scan# 3967
Delta R.T. -0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion	Ratio	Lower	Upper
75	100		
53	105.5	87.6	131.4
89	43.0	37.8	56.8



#82
4-Chlorotoluene
Concen: 52.24 ug/l
RT: 13.106 min Scan# 4145
Delta R.T. 0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion	Ratio	Lower	Upper
91	100		
126	27.3	14.8	44.4









#83
tert-Butylbenzene
Concen: 55.80 ug/l
RT: 13.326 min Scan# 4227
Delta R.T. 0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:119 Resp: 587378
Ion Ratio Lower Upper
119 100
91 79.2 39.5 118.5
134 22.5 12.0 36.0



#84
1,2,4-Trimethylbenzene
Concen: 141.13 ug/l
RT: 13.372 min Scan# 4244
Delta R.T. 0.003 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion:105 Resp: 1845867
Ion Ratio Lower Upper
105 100
120 40.0 20.6 61.8









#85
 sec-Butylbenzene
 Concen: 61.69 ug/l
 RT: 13.503 min Scan# 4293
 Delta R.T. 0.000 min
 Lab File: VN067520.D
 Acq: 25 Jun 2021 21:38




Instrument :
 MSVOA_N
 ClientSampled :


Tgt Ion:105 Resp: 972170
 Ion Ratio Lower Upper
 105 100
 134 17.2 9.0 27.1



#86
 p-Isopropyltoluene
 Concen: 53.38 ug/l
 RT: 13.619 min Scan# 4336
 Delta R.T. 0.001 min
 Lab File: VN067520.D
 Acq: 25 Jun 2021 21:38

Tgt Ion:119 Resp: 625723
 Ion Ratio Lower Upper
 119 100
 134 23.7 12.4 37.4
 91 28.8 14.8 44.3









#87
1,3-Dichlorobenzene
Concen: 49.50 ug/l
RT: 13.616 min Scan# 4335
Delta R.T. -0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:146 Resp: 270200
Ion Ratio Lower Upper
146 100
111 51.3 24.3 72.8
148 62.3 32.0 96.0



#88
1,4-Dichlorobenzene
Concen: 48.84 ug/l
RT: 13.696 min Scan# 4365
Delta R.T. 0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion:146 Resp: 266630
Ion Ratio Lower Upper
146 100
111 50.3 23.4 70.2
148 63.9 31.8 95.3









#89
n-Butylbenzene
Concen: 60.23 ug/l
RT: 13.943 min Scan# 4457
Delta R.T. 0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 91 Resp: 707269
Ion Ratio Lower Upper
91 100
92 49.7 25.7 77.1
134 21.4 11.2 33.6



#90
Hexachloroethane
Concen: 58.98 ug/l
RT: 14.214 min Scan# 4558
Delta R.T. 0.003 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion: 117 Resp: 131757
Ion Ratio Lower Upper
117 100
201 47.5 33.5 100.5









#91
1,2-Dichlorobenzene
Concen: 51.25 ug/l
RT: 13.989 min Scan# 4474
Delta R.T. -0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:146 Resp: 273217
Ion Ratio Lower Upper
146 100
111 53.6 24.9 74.6
148 63.4 31.6 95.0



#92
1,2-Dibromo-3-Chloropropane
Concen: 59.66 ug/l
RT: 14.603 min Scan# 4703
Delta R.T. -0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion: 75 Resp: 56130
Ion Ratio Lower Upper
75 100
155 52.4 29.8 89.5
157 65.9 37.6 112.8









#93
1,2,4-Trichlorobenzene
Concen: 44.07 ug/l
RT: 15.265 min Scan# 4950
Delta R.T. 0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:180 Resp: 117992
Ion Ratio Lower Upper
180 100
182 95.9 47.5 142.5
145 39.3 17.2 51.6



#94
Hexachlorobutadiene
Concen: 43.29 ug/l
RT: 15.373 min Scan# 4990
Delta R.T. -0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion:225 Resp: 52266
Ion Ratio Lower Upper
225 100
223 65.8 31.4 94.0
227 62.7 31.2 93.6









#95
Naphthalene
Concen: 46.83 ug/l
RT: 15.507 min Scan# 5040
Delta R.T. -0.000 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Instrument :
MSVOA_N
ClientSampled :

Tgt Ion:128 Resp: 417069
Ion Ratio Lower Upper
128 100
127 14.0 10.6 15.8
129 11.1 8.6 12.8



#96
1,2,3-Trichlorobenzene
Concen: 45.16 ug/l
RT: 15.703 min Scan# 5113
Delta R.T. 0.001 min
Lab File: VN067520.D
Acq: 25 Jun 2021 21:38

Tgt Ion:180 Resp: 115267
Ion Ratio Lower Upper
180 100
182 91.2 47.8 143.3
145 38.3 18.2 54.6

