

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062521\
 Data File : VN067502.D
 Acq On : 25 Jun 2021 14:18
 Operator : JC/MD
 Sample : M2806-09
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Jun 29 02:00:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062421W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 24 12:18:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	201503	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	457291	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	412603	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	132202	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	232014	56.260	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	112.520%
35) Dibromofluoromethane	8.024	113	141089	50.080	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.160%
50) Toluene-d8	10.443	98	639393	53.212	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.420%
62) 4-Bromofluorobenzene	12.733	95	214392	50.079	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062521\
Data File : VN067502.D
Acq On : 25 Jun 2021 14:18
Operator : JC/MD
Sample : M2806-09
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Jun 29 02:00:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062421W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 24 12:18:45 2021
Response via : Initial Calibration

