

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062521\
 Data File : VN067503.D
 Acq On : 25 Jun 2021 14:42
 Operator : JC/MD
 Sample : M2806-01 10X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1

Manual Integrations
 APPROVED

SAM
 6/29/2021 7:24:09 PM

Quant Time: Jun 29 02:00:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062421W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 24 12:18:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	207914	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	469554	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	430180	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	146613	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	237430	55.798	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 111.600%			
35) Dibromofluoromethane	8.024	113	147018	50.822	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 101.640%			
50) Toluene-d8	10.443	98	653256	52.946	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 105.900%			
62) 4-Bromofluorobenzene	12.733	95	228546	51.991	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 103.980%			
Target Compounds						Qvalue
11) Tert butyl alcohol	5.396	59	6540m	9.411	ug/l	
67) Ethyl Benzene	11.846	91	18128	0.995	ug/l	97
68) m/p-Xylenes	11.961	106	5965	0.968	ug/l	91
73) Isopropylbenzene	12.583	105	458918	30.748	ug/l	98
78) n-propylbenzene	12.924	91	937324	50.929	ug/l	95
83) tert-Butylbenzene	13.326	119	35909	3.731	ug/l	95
84) 1,2,4-Trimethylbenzene	13.372	105	1012447	84.670	ug/l	99
85) sec-Butylbenzene	13.503	105	162105	11.252	ug/l	84
89) n-Butylbenzene	13.946	91	58625	5.461	ug/l #	77
95) Naphthalene	15.512	128	22657	5.921	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062521\
Data File : VN067503.D
Acq On : 25 Jun 2021 14:42
Operator : JC/MD
Sample : M2806-01 10X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1

Manual Integrations
APPROVED

SAM
6/29/2021 7:24:09 PM

Quant Time: Jun 29 02:00:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062421W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 24 12:18:45 2021
Response via : Initial Calibration

