

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062521\
 Data File : VN067509.D
 Acq On : 25 Jun 2021 17:09
 Operator : JC/MD
 Sample : M2827-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ERM-7

Quant Time: Jun 29 02:02:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062421W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 24 12:18:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	195385	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	450112	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	404784	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	128566	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	226540	56.653	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	113.300%
35) Dibromofluoromethane	8.026	113	138253	49.856	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.720%
50) Toluene-d8	10.443	98	624243	52.780	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.560%
62) 4-Bromofluorobenzene	12.736	95	210236	49.892	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062521\
Data File : VN067509.D
Acq On : 25 Jun 2021 17:09
Operator : JC/MD
Sample : M2827-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ERM-7

Quant Time: Jun 29 02:02:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062421W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 24 12:18:45 2021
Response via : Initial Calibration

