

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062521\
 Data File : VN067510.D
 Acq On : 25 Jun 2021 17:33
 Operator : JC/MD
 Sample : M2827-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DAY-5

Quant Time: Jun 29 02:02:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062421W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 24 12:18:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	192535	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	432969	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	385265	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	125044	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	216769	55.012	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	110.020%
35) Dibromofluoromethane	8.024	113	135385	50.755	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.520%
50) Toluene-d8	10.446	98	597172	52.490	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.980%
62) 4-Bromofluorobenzene	12.734	95	200748	49.526	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062521\
Data File : VN067510.D
Acq On : 25 Jun 2021 17:33
Operator : JC/MD
Sample : M2827-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DAY-5

Quant Time: Jun 29 02:02:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062421W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 24 12:18:45 2021
Response via : Initial Calibration

