

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062521\
 Data File : VN067514.D
 Acq On : 25 Jun 2021 19:11
 Operator : JC/MD
 Sample : M2847-03 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1970-WATER

Quant Time: Jun 29 02:02:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062421W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 24 12:18:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	193376	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	437440	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	394531	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	129215	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	222144	56.131	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	112.260%
35) Dibromofluoromethane	8.024	113	137280	50.940	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.880%
50) Toluene-d8	10.443	98	607599	52.861	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.720%
62) 4-Bromofluorobenzene	12.734	95	207455	50.658	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.320%
Target Compounds						
					Qvalue	
16) Acetone	4.299	43	5646	3.727	ug/l #	88
40) Benzene	8.464	78	6884	0.476	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062521\
Data File : VN067514.D
Acq On : 25 Jun 2021 19:11
Operator : JC/MD
Sample : M2847-03 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1970-WATER

Quant Time: Jun 29 02:02:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062421W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 24 12:18:45 2021
Response via : Initial Calibration

