

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062521\
 Data File : VN067516.D
 Acq On : 25 Jun 2021 20:00
 Operator : JC/MD
 Sample : M2812-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 KG-WATER

Quant Time: Jun 29 02:03:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062421W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 24 12:18:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	187039	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	419611	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	365797	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.680	152	109690	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	211019	55.126	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 110.260%	
35) Dibromofluoromethane	8.027	113	131918	51.030	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 102.060%	
50) Toluene-d8	10.446	98	578194	52.440	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 104.880%	
62) 4-Bromofluorobenzene	12.733	95	186016	47.353	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 94.700%	
Target Compounds						
16) Acetone	4.299	43	18370	12.539	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062521\
Data File : VN067516.D
Acq On : 25 Jun 2021 20:00
Operator : JC/MD
Sample : M2812-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
KG-WATER

Quant Time: Jun 29 02:03:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062421W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 24 12:18:45 2021
Response via : Initial Calibration

