

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062521\
 Data File : VN067502.D
 Acq On : 25 Jun 2021 14:18
 Operator : JC/MD
 Sample : M2806-09
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Jun 29 02:00:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062421W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 24 12:18:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	201503	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	457291	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	412603	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	132202	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	232014	56.260	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	112.520%
35) Dibromofluoromethane	8.024	113	141089	50.080	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.160%
50) Toluene-d8	10.443	98	639393	53.212	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.420%
62) 4-Bromofluorobenzene	12.733	95	214392	50.079	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.160%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062521\
Data File : VN067502.D
Acq On : 25 Jun 2021 14:18
Operator : JC/MD
Sample : M2806-09
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Jun 29 02:00:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062421W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 24 12:18:45 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.091 min Scan# 2275
Delta R.T. 0.005 min
Lab File: VN067502.D
Acq: 25 Jun 2021 14:18




Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK


Tgt Ion:168 Resp: 201503
Ion Ratio Lower Upper
168 100
99 88.2 62.9 94.3



#33
1,2-Dichloroethane-d4
Concen: 56.260 ug/l
RT: 8.440 min Scan# 2405
Delta R.T. -0.000 min
Lab File: VN067502.D
Acq: 25 Jun 2021 14:18

Tgt Ion: 65 Resp: 232014
Ion Ratio Lower Upper
65 100
67 52.7 0.0 98.6









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.971 min Scan# 2603
Delta R.T. -0.000 min
Lab File: VN067502.D
Acq: 25 Jun 2021 14:18

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Tgt Ion:114 Resp: 457291




Ion	Ratio	Lower	Upper
114	100		
63	27.8	0.0	49.0
88	18.2	0.0	36.6




#35
Dibromofluoromethane
Concen: 50.080 ug/l
RT: 8.024 min Scan# 2250
Delta R.T. -0.000 min
Lab File: VN067502.D
Acq: 25 Jun 2021 14:18

Tgt Ion:113 Resp: 141089

Ion	Ratio	Lower	Upper
113	100		
111	105.8	83.8	125.8
192	12.6	13.4	20.0#









#50
Toluene-d8
Concen: 53.212 ug/l
RT: 10.443 min Scan# 3152
Delta R.T. -0.000 min
Lab File: VN067502.D
Acq: 25 Jun 2021 14:18




Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK


Tgt Ion: 98 Resp: 639393
Ion Ratio Lower Upper
98 100
100 61.3 51.7 77.5



#62
4-Bromofluorobenzene
Concen: 50.079 ug/l
RT: 12.733 min Scan# 4006
Delta R.T. 0.000 min
Lab File: VN067502.D
Acq: 25 Jun 2021 14:18

Tgt Ion: 95 Resp: 214392
Ion Ratio Lower Upper
95 100
174 50.4 0.0 122.0
176 48.7 0.0 117.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.749 min Scan# 3639
Delta R.T. 0.003 min
Lab File: VN067502.D
Acq: 25 Jun 2021 14:18

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Tgt Ion:117 Resp: 412603
Ion Ratio Lower Upper
117 100
82 70.3 53.1 79.7
119 31.8 25.5 38.3



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.677 min Scan# 4358
Delta R.T. 0.002 min
Lab File: VN067502.D
Acq: 25 Jun 2021 14:18

Tgt Ion:152 Resp: 132202
Ion Ratio Lower Upper
152 100
115 77.6 36.1 108.5
150 159.1 0.0 351.8

