

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062521\
 Data File : VN067512.D
 Acq On : 25 Jun 2021 18:22
 Operator : JC/MD
 Sample : M2847-04 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1962-1964

Quant Time: Jun 29 02:02:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062421W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 24 12:18:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	191526	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	432672	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	385164	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	119037	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	217281	55.432	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	110.860%
35) Dibromofluoromethane	8.024	113	131449	49.313	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.620%
50) Toluene-d8	10.443	98	591930	52.065	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.140%
62) 4-Bromofluorobenzene	12.734	95	197626	48.789	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.580%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062521\
Data File : VN067512.D
Acq On : 25 Jun 2021 18:22
Operator : JC/MD
Sample : M2847-04 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1962-1964

Quant Time: Jun 29 02:02:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062421W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 24 12:18:45 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.088 min Scan# 2274
Delta R.T. 0.002 min
Lab File: VN067512.D
Acq: 25 Jun 2021 18:22




Instrument :
MSVOA_N
ClientSampled :
1962-1964


Tgt Ion:168 Resp: 191526
Ion Ratio Lower Upper
168 100
99 88.0 62.9 94.3



#33
1,2-Dichloroethane-d4
Concen: 55.432 ug/l
RT: 8.440 min Scan# 2405
Delta R.T. -0.000 min
Lab File: VN067512.D
Acq: 25 Jun 2021 18:22

Tgt Ion: 65 Resp: 217281
Ion Ratio Lower Upper
65 100
67 53.6 0.0 98.6









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.971 min Scan# 2603
Delta R.T. -0.000 min
Lab File: VN067512.D
Acq: 25 Jun 2021 18:22

Instrument :
MSVOA_N
ClientSampled :
1962-1964

Tgt Ion:114 Resp: 432672




Ion	Ratio	Lower	Upper
114	100		
63	27.6	0.0	49.0
88	18.8	0.0	36.6




#35
Dibromofluoromethane
Concen: 49.313 ug/l
RT: 8.024 min Scan# 2250
Delta R.T. -0.000 min
Lab File: VN067512.D
Acq: 25 Jun 2021 18:22

Tgt Ion:113 Resp: 131449

Ion	Ratio	Lower	Upper
113	100		
111	105.5	83.8	125.8
192	12.4	13.4	20.0#









#50
Toluene-d8
Concen: 52.065 ug/l
RT: 10.443 min Scan# 3152
Delta R.T. 0.000 min
Lab File: VN067512.D
Acq: 25 Jun 2021 18:22




Instrument :
MSVOA_N
ClientSampled :
1962-1964


Tgt Ion: 98 Resp: 591930
Ion Ratio Lower Upper
98 100
100 63.3 51.7 77.5



#62
4-Bromofluorobenzene
Concen: 48.789 ug/l
RT: 12.734 min Scan# 4006
Delta R.T. 0.001 min
Lab File: VN067512.D
Acq: 25 Jun 2021 18:22

Tgt Ion: 95 Resp: 197626
Ion Ratio Lower Upper
95 100
174 50.5 0.0 122.0
176 47.5 0.0 117.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.749 min Scan# 3639
Delta R.T. 0.003 min
Lab File: VN067512.D
Acq: 25 Jun 2021 18:22

Instrument :
MSVOA_N
ClientSampleId :
1962-1964

Tgt Ion:117 Resp: 385164
Ion Ratio Lower Upper
117 100
82 68.4 53.1 79.7
119 32.4 25.5 38.3



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.678 min Scan# 4358
Delta R.T. 0.003 min
Lab File: VN067512.D
Acq: 25 Jun 2021 18:22

Tgt Ion:152 Resp: 119037
Ion Ratio Lower Upper
152 100
115 76.7 36.1 108.5
150 159.0 0.0 351.8

