



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062521\
 Data File : VN067497.D
 Acq On : 25 Jun 2021 11:02
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062421W.M
 Title : SW846 8260
 Last Update : Thu Jun 24 12:18:45 2021



AutoFind: Scans 4005, 4006, 4007; Background Corrected with Scan 3989

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.1	26568	PASS
75	95	30	60	55.9	70544	PASS
95	95	100	100	100.0	126117	PASS
96	95	5	9	6.4	8040	PASS
173	174	0.00	2	1.0	923	PASS
174	95	50	100	74.8	94312	PASS
175	174	5	9	8.2	7701	PASS
176	174	95	101	96.6	91144	PASS
177	176	5	9	6.5	5910	PASS