

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062522\
 Data File : VN073105.D
 Acq On : 25 Jun 2022 13:02
 Operator : JC\MD
 Sample : N3431-15
 Misc : 5.63g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-14B-2-4

Quant Time: Jun 27 03:37:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.010	168	336849	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	545721	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	510399	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	370593	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.363	65	189650	41.644	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	83.280%		
35) Dibromofluoromethane	7.945	113	159587	47.312	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.620%		
50) Toluene-d8	10.374	98	826137	65.821	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	131.640%#		
62) 4-Bromofluorobenzene	12.668	95	487666	108.052	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	216.100%#		
Target Compounds						
					Qvalue	
16) Acetone	4.251	43	8151	3.256	ug/l #	87
18) Methyl Acetate	4.775	43	31617	4.743	ug/l	91
31) Cyclohexane	8.004	56	65148	10.575	ug/l #	91
39) Methylcyclohexane	9.386	83	1476919	268.465	ug/l	91
40) Benzene	8.375	78	27419	1.912	ug/l #	89
73) Isopropylbenzene	12.521	105	963153	33.594	ug/l	99
78) n-propylbenzene	12.857	91	930582	29.509	ug/l	96
83) tert-Butylbenzene	13.263	119	173536	8.046	ug/l	76
84) 1,2,4-Trimethylbenzene	13.304	105	113168	4.825	ug/l	82
85) sec-Butylbenzene	13.439	105	1328568	46.176	ug/l	80
89) n-Butylbenzene	13.880	91	629714	33.107	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062522\
Data File : VN073105.D
Acq On : 25 Jun 2022 13:02
Operator : JC\MD
Sample : N3431-15
Misc : 5.63g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-14B-2-4

Quant Time: Jun 27 03:37:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

