

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062522\
 Data File : VN073111.D
 Acq On : 25 Jun 2022 15:29
 Operator : JC\MD
 Sample : N3454-01
 Misc : 4.83g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-16-1A-1-GR

Quant Time: Jun 27 03:41:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.010	168	359048	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	560073	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	520500	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	292587	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.363	65	198834	40.961	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	81.920%		
35) Dibromofluoromethane	7.945	113	169368	48.925	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.860%		
50) Toluene-d8	10.374	98	695908	54.025	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	108.040%		
62) 4-Bromofluorobenzene	12.668	95	347963	75.123	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	150.240%#		
Target Compounds						Qvalue
16) Acetone	4.257	43	13422	5.030	ug/l	90
18) Methyl Acetate	4.775	43	30096	4.236	ug/l	93
25) 2-Butanone	7.286	43	17316	4.090	ug/l	89
31) Cyclohexane	8.004	56	351194	53.484	ug/l	93
39) Methylcyclohexane	9.386	83	1513289	268.027	ug/l	93
40) Benzene	8.375	78	42475	2.885	ug/l #	90
52) Toluene	10.439	92	30676	3.394	ug/l	90
67) Ethyl Benzene	11.780	91	76362	4.267	ug/l	98
68) m/p-Xylenes	11.886	106	102387	14.588	ug/l	92
69) o-Xylene	12.216	106	36950	5.269	ug/l	74
73) Isopropylbenzene	12.516	105	398349	17.599	ug/l	98
78) n-propylbenzene	12.857	91	333125	13.380	ug/l	97
80) 1,3,5-Trimethylbenzene	12.998	105	131253	7.059	ug/l	90
83) tert-Butylbenzene	13.263	119	84255	4.948	ug/l	88
84) 1,2,4-Trimethylbenzene	13.304	105	453176	24.471	ug/l	95
85) sec-Butylbenzene	13.439	105	435796	19.185	ug/l #	76
86) p-Isopropyltoluene	13.557	119	92041	5.005	ug/l	89
89) n-Butylbenzene	13.880	91	355628	23.682	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062522\
Data File : VN073111.D
Acq On : 25 Jun 2022 15:29
Operator : JC\MD
Sample : N3454-01
Misc : 4.83g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-16-1A-1-GR

Quant Time: Jun 27 03:41:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

