

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062522\
 Data File : VN073120.D
 Acq On : 25 Jun 2022 19:10
 Operator : JC\MD
 Sample : N3434-03 40X
 Misc : 5.65g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 B2204-27.5

Quant Time: Jun 27 03:46:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	365221	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	568026	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	537577	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	273064	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	213488	43.236	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.480%
35) Dibromofluoromethane	7.951	113	176179	50.180	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.360%
50) Toluene-d8	10.380	98	601327	46.029	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.060%
62) 4-Bromofluorobenzene	12.668	95	272870	58.086	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	116.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062522\
Data File : VN073120.D
Acq On : 25 Jun 2022 19:10
Operator : JC\MD
Sample : N3434-03 40X
Misc : 5.65g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
B2204-27.5

Quant Time: Jun 27 03:46:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

