

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062522\
 Data File : VN073124.D
 Acq On : 25 Jun 2022 20:48
 Operator : JC\MD
 Sample : N3434-07 40X
 Misc : 6.88g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 B2206-29

Quant Time: Jun 27 03:47:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	354940	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	559592	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	526168	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	272179	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	206663	43.066	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	86.140%		
35) Dibromofluoromethane	7.951	113	175861	50.845	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	101.680%		
50) Toluene-d8	10.375	98	591079	45.926	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	91.860%		
62) 4-Bromofluorobenzene	12.669	95	270975	58.552	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	117.100%		
Target Compounds						Qvalue
39) Methylcyclohexane	9.392	83	16434	2.913	ug/l #	83
52) Toluene	10.445	92	18283	2.025	ug/l	97
67) Ethyl Benzene	11.780	91	118065	6.527	ug/l	97
68) m/p-Xylenes	11.886	106	189218	26.670	ug/l	96
69) o-Xylene	12.216	106	66701	9.408	ug/l	95
78) n-propylbenzene	12.857	91	57106	2.466	ug/l	98
80) 1,3,5-Trimethylbenzene	12.998	105	104987	6.070	ug/l	97
84) 1,2,4-Trimethylbenzene	13.304	105	311071	18.057	ug/l	99
95) Naphthalene	15.433	128	49540	2.356	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062522\
Data File : VN073124.D
Acq On : 25 Jun 2022 20:48
Operator : JC\MD
Sample : N3434-07 40X
Misc : 6.88g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
B2206-29

Quant Time: Jun 27 03:47:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

