

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062522\
 Data File : VN073128.D
 Acq On : 26 Jun 2022 00:28
 Operator : JC\MD
 Sample : VN0625MBL02
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0625MBL02

Quant Time: Jun 27 03:50:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	326501	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	515710	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	465730	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	218880	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	194016	43.953	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.900%
35) Dibromofluoromethane	7.951	113	164129	51.491	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.980%
50) Toluene-d8	10.381	98	539615	45.495	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.000%
62) 4-Bromofluorobenzene	12.669	95	232593	54.535	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062522\
Data File : VN073128.D
Acq On : 26 Jun 2022 00:28
Operator : JC\MD
Sample : VN0625MBL02
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0625MBL02

Quant Time: Jun 27 03:50:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

