

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062522\
 Data File : VN073132.D
 Acq On : 26 Jun 2022 02:05
 Operator : JC\MD
 Sample : N3494-06
 Misc : 4.99g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP-EXC-1-4

Quant Time: Jun 27 03:52:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.010	168	327778	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	522152	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	519914	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	281294	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.363	65	188228	42.475	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	84.960%
35) Dibromofluoromethane	7.945	113	157410	48.773	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.540%
50) Toluene-d8	10.375	98	562599	46.848	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.700%
62) 4-Bromofluorobenzene	12.669	95	275445	63.785	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	127.580%#

Target Compounds						Qvalue
18) Methyl Acetate	4.775	43	262496	40.466	ug/l	91
40) Benzene	8.375	78	22294	1.624	ug/l #	85
52) Toluene	10.439	92	15397	1.827	ug/l	100
65) Chlorobenzene	11.710	112	163684	16.121	ug/l	98
67) Ethyl Benzene	11.780	91	27214	1.523	ug/l	89
68) m/p-Xylenes	11.886	106	30721	4.382	ug/l	92
78) n-propylbenzene	12.857	91	28560	1.193	ug/l	98
84) 1,2,4-Trimethylbenzene	13.304	105	30172	1.695	ug/l	81
91) 1,2-Dichlorobenzene	13.921	146	29539	3.043	ug/l #	72
95) Naphthalene	15.433	128	59786	2.751	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062522\
Data File : VN073132.D
Acq On : 26 Jun 2022 02:05
Operator : JC\MD
Sample : N3494-06
Misc : 4.99g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP-EXC-1-4

Quant Time: Jun 27 03:52:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

