

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062522\
 Data File : VN073136.D
 Acq On : 26 Jun 2022 03:43
 Operator : JC\MD
 Sample : N3494-26
 Misc : 4.97g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP-EXC-1-2

Quant Time: Jun 27 03:55:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.010	168	355086	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	560053	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	546132	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	389238	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.363	65	200827	41.833	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	83.660%
35) Dibromofluoromethane	7.945	113	169220	48.884	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.760%
50) Toluene-d8	10.374	98	605247	46.988	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.980%
62) 4-Bromofluorobenzene	12.668	95	302001	65.202	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	130.400%#
Target Compounds						
					Qvalue	
18) Methyl Acetate	4.781	43	12297	1.750	ug/l #	88
40) Benzene	8.381	78	76838	5.220	ug/l	96
52) Toluene	10.439	92	36909	4.084	ug/l	98
65) Chlorobenzene	11.704	112	160951	15.091	ug/l	96
67) Ethyl Benzene	11.780	91	55143	2.937	ug/l	97
68) m/p-Xylenes	11.892	106	52075	7.072	ug/l	98
78) n-propylbenzene	12.857	91	85275	2.575	ug/l	99
84) 1,2,4-Trimethylbenzene	13.304	105	49946	2.027	ug/l	87
95) Naphthalene	15.433	128	1170333	38.921	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062522\
Data File : VN073136.D
Acq On : 26 Jun 2022 03:43
Operator : JC\MD
Sample : N3494-26
Misc : 4.97g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP-EXC-1-2

Quant Time: Jun 27 03:55:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

