

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062522\
 Data File : VN073138.D
 Acq On : 26 Jun 2022 04:32
 Operator : JC\MD
 Sample : N3454-04 20X
 Misc : 3.80g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-15-1-2-GR

Quant Time: Jun 27 03:56:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	392379	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	620692	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	581967	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.609	152	304680	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	228096	42.998	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.000%
35) Dibromofluoromethane	7.951	113	190289	49.600	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.200%
50) Toluene-d8	10.374	98	657060	46.027	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.060%
62) 4-Bromofluorobenzene	12.668	95	306680	59.744	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	119.480%
Target Compounds						
18) Methyl Acetate	4.781	43	100206	12.904	ug/l #	88
95) Naphthalene	15.433	128	145104	6.165	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062522\
Data File : VN073138.D
Acq On : 26 Jun 2022 04:32
Operator : JC\MD
Sample : N3454-04 20X
Misc : 3.80g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-15-1-2-GR

Quant Time: Jun 27 03:56:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

