

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062522\
 Data File : VN073144.D
 Acq On : 26 Jun 2022 06:58
 Operator : JC\MD
 Sample : N3426-04 40X
 Misc : 6.11g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 B2202-30

Quant Time: Jun 27 03:58:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	366111	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	582864	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	550556	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	286807	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	214433	43.322	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	86.640%		
35) Dibromofluoromethane	7.951	113	178912	49.661	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.320%		
50) Toluene-d8	10.380	98	617641	46.074	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	92.140%		
62) 4-Bromofluorobenzene	12.669	95	281099	58.314	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	116.620%		
Target Compounds						Qvalue
39) Methylcyclohexane	9.392	83	32696	5.565	ug/l #	81
52) Toluene	10.439	92	26214	2.787	ug/l	95
67) Ethyl Benzene	11.780	91	269935	14.262	ug/l	98
68) m/p-Xylenes	11.886	106	409448	55.154	ug/l	96
69) o-Xylene	12.216	106	102039	13.756	ug/l	98
73) Isopropylbenzene	12.516	105	30844	1.390	ug/l	97
78) n-propylbenzene	12.857	91	131189	5.375	ug/l	99
80) 1,3,5-Trimethylbenzene	12.998	105	211893	11.626	ug/l	98
84) 1,2,4-Trimethylbenzene	13.304	105	669981	36.907	ug/l	97
89) n-Butylbenzene	13.880	91	19763	1.343	ug/l #	67
95) Naphthalene	15.433	128	125478	5.663	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062522\
Data File : VN073144.D
Acq On : 26 Jun 2022 06:58
Operator : JC\MD
Sample : N3426-04 40X
Misc : 6.11g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
B2202-30

Quant Time: Jun 27 03:58:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

