

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062522\
 Data File : VN073145.D
 Acq On : 26 Jun 2022 07:22
 Operator : JC\MD
 Sample : N3426-05 40X
 Misc : 6.89g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 B2202-35

Quant Time: Jun 27 03:59:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	350130	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	564319	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	525466	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	262689	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	204744	43.253	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.500%
35) Dibromofluoromethane	7.951	113	176927	50.724	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.440%
50) Toluene-d8	10.374	98	592129	45.622	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.240%
62) 4-Bromofluorobenzene	12.668	95	268867	57.610	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	115.220%
Target Compounds						
95) Naphthalene	15.427	128	21984	1.083	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062522\
Data File : VN073145.D
Acq On : 26 Jun 2022 07:22
Operator : JC\MD
Sample : N3426-05 40X
Misc : 6.89g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
B2202-35

Quant Time: Jun 27 03:59:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

