

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062524\
 Data File : VN082705.D
 Acq On : 25 Jun 2024 09:56
 Operator : JC\MD
 Sample : VN0625WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0625WBL01

Quant Time: Jun 26 01:17:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	151943	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	276953	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	270778	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	111806	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	124647	51.156	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.320%
35) Dibromofluoromethane	8.165	113	93603	49.820	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.640%
50) Toluene-d8	10.571	98	332840	46.846	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.700%
62) 4-Bromofluorobenzene	12.853	95	105715	43.157	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	86.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062524\
Data File : VN082705.D
Acq On : 25 Jun 2024 09:56
Operator : JC\MD
Sample : VN0625WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0625WBL01

Quant Time: Jun 26 01:17:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

