

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062524\
 Data File : VN082709.D
 Acq On : 25 Jun 2024 11:43
 Operator : JC\MD
 Sample : PB161676ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161676ZHE#01

Quant Time: Jun 26 01:19:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	140806	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	265810	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	282483	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	114192	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	126846	56.176	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.360%
35) Dibromofluoromethane	8.165	113	93541	51.874	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.740%
50) Toluene-d8	10.571	98	363699	53.335	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.660%
62) 4-Bromofluorobenzene	12.853	95	118247	50.296	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	100.600%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	31243	30.424	ug/l	99
20) Methylene Chloride	5.283	84	8612	4.770	ug/l	97
43) Isopropyl Acetate	8.688	43	203781	31.756	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062524\
Data File : VN082709.D
Acq On : 25 Jun 2024 11:43
Operator : JC\MD
Sample : PB161676ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB161676ZHE#01

Quant Time: Jun 26 01:19:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

