

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062524\
 Data File : VN082710.D
 Acq On : 25 Jun 2024 12:07
 Operator : JC\MD
 Sample : PB161676ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161676ZHE#02

Quant Time: Jun 26 01:20:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	135315	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	261934	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	272996	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	108107	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	127006	58.529	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	117.060%
35) Dibromofluoromethane	8.165	113	93710	52.737	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.480%
50) Toluene-d8	10.571	98	358169	53.301	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.600%
62) 4-Bromofluorobenzene	12.853	95	113212	48.867	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	97.740%
Target Compounds						
						Qvalue
16) Acetone	4.436	43	31974	32.399	ug/l	98
20) Methylene Chloride	5.283	84	9023	5.201	ug/l #	82
43) Isopropyl Acetate	8.688	43	215785	34.124	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062524\
Data File : VN082710.D
Acq On : 25 Jun 2024 12:07
Operator : JC\MD
Sample : PB161676ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB161676ZHE#02

Quant Time: Jun 26 01:20:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

