

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062524\
 Data File : VN082711.D
 Acq On : 25 Jun 2024 12:31
 Operator : JC\MD
 Sample : PB161676ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161676ZHE#03

Quant Time: Jun 26 01:20:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	141531	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	275574	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	289991	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	113252	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	129896	57.232	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.460%
35) Dibromofluoromethane	8.165	113	98348	52.607	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.220%
50) Toluene-d8	10.571	98	382742	54.139	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.280%
62) 4-Bromofluorobenzene	12.847	95	121211	49.730	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	99.460%
Target Compounds						
						Qvalue
16) Acetone	4.436	43	28878	27.977	ug/l	96
20) Methylene Chloride	5.271	84	9519	5.245	ug/l #	81
43) Isopropyl Acetate	8.694	43	236358	35.527	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062524\
 Data File : VN082711.D
 Acq On : 25 Jun 2024 12:31
 Operator : JC\MD
 Sample : PB161676ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161676ZHE#03

Quant Time: Jun 26 01:20:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

