

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062524\
 Data File : VN082712.D
 Acq On : 25 Jun 2024 12:56
 Operator : JC\MD
 Sample : PB161676ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161676ZHE#04

Quant Time: Jun 26 01:21:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	133635	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	262742	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	279250	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	110677	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	126469	59.014	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 118.020%		
35) Dibromofluoromethane	8.171	113	94405	52.965	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 105.920%		
50) Toluene-d8	10.571	98	362070	53.716	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 107.440%		
62) 4-Bromofluorobenzene	12.853	95	113197	48.711	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 97.420%		
Target Compounds						
						Qvalue
16) Acetone	4.436	43	31471	32.291	ug/l	90
20) Methylene Chloride	5.277	84	9369	5.468	ug/l #	88
43) Isopropyl Acetate	8.688	43	234799	37.017	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062524\
Data File : VN082712.D
Acq On : 25 Jun 2024 12:56
Operator : JC\MD
Sample : PB161676ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB161676ZHE#04

Quant Time: Jun 26 01:21:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

