

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062524\
 Data File : VN082713.D
 Acq On : 25 Jun 2024 13:20
 Operator : JC\MD
 Sample : PB161676ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161676ZHE#05

Quant Time: Jun 26 01:21:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	132990	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	263214	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	278428	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	109698	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	124285	58.277	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.560%
35) Dibromofluoromethane	8.171	113	93455	52.338	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.680%
50) Toluene-d8	10.571	98	359811	53.285	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.580%
62) 4-Bromofluorobenzene	12.853	95	116333	49.970	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	99.940%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	32683	33.697	ug/l	99
20) Methylene Chloride	5.283	84	8659	5.078	ug/l	98
43) Isopropyl Acetate	8.694	43	274296	43.166	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062524\
Data File : VN082713.D
Acq On : 25 Jun 2024 13:20
Operator : JC\MD
Sample : PB161676ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB161676ZHE#05

Quant Time: Jun 26 01:21:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

