

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062524\
 Data File : VN082714.D
 Acq On : 25 Jun 2024 13:44
 Operator : JC\MD
 Sample : PB161676ZHE#06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161676ZHE#06

Quant Time: Jun 26 01:21:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	133681	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	264353	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	276108	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	108120	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	126594	59.052	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	118.100%
35) Dibromofluoromethane	8.165	113	94529	52.711	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.420%
50) Toluene-d8	10.570	98	363340	53.576	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.160%
62) 4-Bromofluorobenzene	12.853	95	113667	48.615	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	97.220%
Target Compounds						
						Qvalue
16) Acetone	4.436	43	32670	33.509	ug/l	100
20) Methylene Chloride	5.277	84	8183	4.774	ug/l	83
43) Isopropyl Acetate	8.688	43	276800	43.372	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062524\
Data File : VN082714.D
Acq On : 25 Jun 2024 13:44
Operator : JC\MD
Sample : PB161676ZHE#06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB161676ZHE#06

Quant Time: Jun 26 01:21:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

