

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062524\
 Data File : VN082716.D
 Acq On : 25 Jun 2024 14:33
 Operator : JC\MD
 Sample : PB161676ZHE#08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161676ZHE#08

Quant Time: Jun 26 01:22:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	135474	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	270147	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	281756	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	110365	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	129060	59.406	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 118.820%			
35) Dibromofluoromethane	8.171	113	95626	52.179	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.360%			
50) Toluene-d8	10.571	98	366295	52.853	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 105.700%			
62) 4-Bromofluorobenzene	12.853	95	115203	48.215	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery = 96.420%			
Target Compounds					Qvalue	
16) Acetone	4.436	43	33785	34.194	ug/l	95
20) Methylene Chloride	5.277	84	8770	5.049	ug/l	88
43) Isopropyl Acetate	8.694	43	313196	48.023	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062524\
Data File : VN082716.D
Acq On : 25 Jun 2024 14:33
Operator : JC\MD
Sample : PB161676ZHE#08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB161676ZHE#08

Quant Time: Jun 26 01:22:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

