

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062524\
 Data File : VN082719.D
 Acq On : 25 Jun 2024 15:46
 Operator : JC\MD
 Sample : PB161676ZHE#11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161676ZHE#11

Quant Time: Jun 26 01:24:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	122430	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	241508	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	249101	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	93444	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	116873	59.528	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 119.060%	
35) Dibromofluoromethane	8.171	113	87590	53.462	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.920%	
50) Toluene-d8	10.571	98	333019	53.750	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.500%	
62) 4-Bromofluorobenzene	12.853	95	101414	47.477	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 94.960%	
Target Compounds						Qvalue
16) Acetone	4.436	43	33330	37.328	ug/l	92
20) Methylene Chloride	5.277	84	11835	7.539	ug/l	91
43) Isopropyl Acetate	8.694	43	245795	42.157	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062524\
Data File : VN082719.D
Acq On : 25 Jun 2024 15:46
Operator : JC\MD
Sample : PB161676ZHE#11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB161676ZHE#11

Quant Time: Jun 26 01:24:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

