

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN062524\  
 Data File : VN082721.D  
 Acq On : 25 Jun 2024 16:35  
 Operator : JC\MD  
 Sample : P3008-03  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 20 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 WC-WS-6-21-24

Quant Time: Jun 26 01:25:02 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N061724W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Jun 18 04:52:44 2024  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) Pentafluorobenzene	8.229	168	130929	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	253994	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	265899	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	102968	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	125094	59.579	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	119.160%
35) Dibromofluoromethane	8.165	113	91960	53.370	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.740%
50) Toluene-d8	10.570	98	347866	53.386	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.780%
62) 4-Bromofluorobenzene	12.853	95	106868	47.571	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	95.140%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	65592	68.691	ug/l	97
20) Methylene Chloride	5.283	84	7718	4.597	ug/l #	70
25) 2-Butanone	7.488	43	16098	11.861	ug/l	98
43) Isopropyl Acetate	8.694	43	197227	32.164	ug/l #	83
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN062524\  
 Data File : VN082721.D  
 Acq On : 25 Jun 2024 16:35  
 Operator : JC\MD  
 Sample : P3008-03  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 20 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 WC-WS-6-21-24

Quant Time: Jun 26 01:25:02 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N061724W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Jun 18 04:52:44 2024  
 Response via : Initial Calibration

