

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062524\
 Data File : VN082722.D
 Acq On : 25 Jun 2024 17:00
 Operator : JC\MD
 Sample : P3009-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-204

Quant Time: Jun 26 01:25:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	129553	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	246540	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	260329	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	99136	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	122098	58.770	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	117.540%
35) Dibromofluoromethane	8.171	113	92368	55.227	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.460%
50) Toluene-d8	10.570	98	340366	53.814	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.620%
62) 4-Bromofluorobenzene	12.853	95	102505	47.008	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	94.020%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	38097	40.321	ug/l	91
20) Methylene Chloride	5.271	84	7861	4.732	ug/l #	82
43) Isopropyl Acetate	8.694	43	138787	23.318	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062524\
Data File : VN082722.D
Acq On : 25 Jun 2024 17:00
Operator : JC\MD
Sample : P3009-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-204

Quant Time: Jun 26 01:25:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

