

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062524\
 Data File : VN082723.D
 Acq On : 25 Jun 2024 17:24
 Operator : JC\MD
 Sample : P3021-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ARS20-008

Quant Time: Jun 26 01:25:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	131764	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	255522	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	270285	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	105677	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	124107	58.735	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 117.460%	
35) Dibromofluoromethane	8.171	113	93737	54.076	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.160%	
50) Toluene-d8	10.571	98	350452	53.461	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 106.920%	
62) 4-Bromofluorobenzene	12.853	95	106658	47.194	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 94.380%	
Target Compounds						
						Qvalue
16) Acetone	4.442	43	36240	37.712	ug/l	95
20) Methylene Chloride	5.277	84	7651	4.529	ug/l #	78
43) Isopropyl Acetate	8.694	43	167640	27.176	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062524\
Data File : VN082723.D
Acq On : 25 Jun 2024 17:24
Operator : JC\MD
Sample : P3021-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ARS20-008

Quant Time: Jun 26 01:25:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

