

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062524\
 Data File : VN082724.D
 Acq On : 25 Jun 2024 17:49
 Operator : JC\MD
 Sample : P3021-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ARS20-0018

Quant Time: Jun 26 01:26:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	126332	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	246650	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	263878	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	100892	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	121376	59.912	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	119.820%
35) Dibromofluoromethane	8.171	113	91652	54.775	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.540%
50) Toluene-d8	10.570	98	346491	54.758	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.520%
62) 4-Bromofluorobenzene	12.853	95	105375	48.303	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	96.600%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	34010	36.913	ug/l	98
20) Methylene Chloride	5.283	84	8285	5.115	ug/l	89
43) Isopropyl Acetate	8.694	43	168713	28.333	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062524\
Data File : VN082724.D
Acq On : 25 Jun 2024 17:49
Operator : JC\MD
Sample : P3021-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ARS20-0018

Quant Time: Jun 26 01:26:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

