

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062524\
 Data File : VN082725.D
 Acq On : 25 Jun 2024 18:13
 Operator : JC\MD
 Sample : P3021-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ARS20-006

Quant Time: Jun 26 01:26:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	125446	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	243807	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	262059	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	101294	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	120795	60.046	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 120.100%	
35) Dibromofluoromethane	8.171	113	89467	54.092	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.180%	
50) Toluene-d8	10.571	98	342040	54.685	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 109.380%	
62) 4-Bromofluorobenzene	12.853	95	104937	48.663	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 97.320%	
Target Compounds						Qvalue
16) Acetone	4.436	43	33783	36.925	ug/l	99
20) Methylene Chloride	5.277	84	8426	5.239	ug/l	89
43) Isopropyl Acetate	8.694	43	180832	30.723	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062524\
Data File : VN082725.D
Acq On : 25 Jun 2024 18:13
Operator : JC\MD
Sample : P3021-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ARS20-006

Quant Time: Jun 26 01:26:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

