

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062524\
 Data File : VN082726.D
 Acq On : 25 Jun 2024 18:38
 Operator : JC\MD
 Sample : P3021-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ARS20-0015

Quant Time: Jun 26 01:27:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	128052	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	247412	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	259470	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	100848	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	123174	59.983	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	=	119.960%	
35) Dibromofluoromethane	8.171	113	91610	54.581	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	=	109.160%	
50) Toluene-d8	10.570	98	341949	53.874	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	=	107.740%	
62) 4-Bromofluorobenzene	12.853	95	103574	47.331	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery	=	94.660%	
Target Compounds						Qvalue
16) Acetone	4.430	43	33805	36.198	ug/l	93
20) Methylene Chloride	5.277	84	7641	4.654	ug/l #	80
43) Isopropyl Acetate	8.694	43	126047	21.103	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062524\
Data File : VN082726.D
Acq On : 25 Jun 2024 18:38
Operator : JC\MD
Sample : P3021-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ARS20-0015

Quant Time: Jun 26 01:27:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

