

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062618\
 Data File : VN049547.D
 Acq On : 26 Jun 2018 12:36
 Operator : MD\SY
 Sample : VN0626WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0626WBL01

Quant Time: Jun 26 14:53:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1804615	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	2889075	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2438837	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	1063951	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	1175576	49.11	ug/l	0.00
Spiked Amount			Recovery	=	98.22%	
35) Dibromofluoromethane	7.59	113	1042558	44.60	ug/l	0.00
Spiked Amount			Recovery	=	89.20%	
50) Toluene-d8	10.09	98	3943279	44.31	ug/l	0.00
Spiked Amount			Recovery	=	88.62%	
62) 4-Bromofluorobenzene	12.40	95	1212532	41.20	ug/l	0.00
Spiked Amount			Recovery	=	82.40%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	106905	18.40	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062618\
Data File : VN049547.D
Acq On : 26 Jun 2018 12:36
Operator : MD\SY
Sample : VN0626WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0626WBL01

Quant Time: Jun 26 14:53:20 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

