

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062618\
 Data File : VN049552.D
 Acq On : 26 Jun 2018 15:05
 Operator : MD\SY
 Sample : J3241-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HR-06-062218-A

Quant Time: Jun 27 01:24:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1341165	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2134257	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1785936	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	789203	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	860562	48.38	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	96.76%	
35) Dibromofluoromethane	7.59	113	774032	44.83	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	89.66%	
50) Toluene-d8	10.09	98	2897235	44.07	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	88.14%	
62) 4-Bromofluorobenzene	12.40	95	889054	40.90	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	81.80%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	170145	39.39	ug/l	95
18) Methyl Acetate	4.33	43	253283	19.23	ug/l	95
20) Methylene Chloride	4.55	84	1740619	96.49	ug/l	91
44) Trichloroethene	8.84	130	153205	7.74	ug/l	92
64) Tetrachloroethene	10.63	164	18539	1.01	ug/l	98
68) m/p-Xylenes	11.62	106	49576	1.73	ug/l	90
95) Naphthalene	15.13	128	492991	23.27	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062618\
Data File : VN049552.D
Acq On : 26 Jun 2018 15:05
Operator : MD\SY
Sample : J3241-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HR-06-062218-A

Quant Time: Jun 27 01:24:47 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

