

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062619\
 Data File : VN056507.D
 Acq On : 26 Jun 2019 15:34
 Operator : JC/SP
 Sample : K3543-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-TT-MW202D1-GW-20190625

Quant Time: Jun 27 08:44:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 03:45:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	351630	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	602003	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	592499	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	170165	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	202041	51.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.68%	
35) Dibromofluoromethane	7.59	113	177522	47.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.54%	
50) Toluene-d8	10.09	98	748192	51.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.40%	
62) 4-Bromofluorobenzene	12.40	95	236952	48.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.28%	
Target Compounds						
19) Methyl tert-butyl Ether	5.04	73	13929	1.355	ug/l	94
22) Diisopropyl ether	5.95	45	8945	0.824	ug/l #	89
27) cis-1,2-Dichloroethene	6.83	96	68273	17.425	ug/l	90
30) Chloroform	7.37	83	7343	1.194	ug/l	95
40) Benzene	8.04	78	14030	0.861	ug/l	99
44) Trichloroethene	8.83	130	48912	11.140	ug/l	98
64) Tetrachloroethene	10.63	164	70485	15.022	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062619\
Data File : VN056507.D
Acq On : 26 Jun 2019 15:34
Operator : JC/SP
Sample : K3543-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-TT-MW202D1-GW-20190625

Quant Time: Jun 27 08:44:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 03:45:40 2019
Response via : Initial Calibration

