

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062620\
 Data File : VN062133.D
 Acq On : 26 Jun 2020 12:36
 Operator : JC/MD
 Sample : L3043-18DL 10X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE120D1-20200615DL

Quant Time: Jun 27 04:16:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	236164	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	430336	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	404724	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	153451	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	152936	46.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.30%	
35) Dibromofluoromethane	7.55	113	132000	47.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.62%	
50) Toluene-d8	10.06	98	525491	47.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.24%	
62) 4-Bromofluorobenzene	12.38	95	174273	46.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.48%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	3.71	101	5172	1.907	ug/l #	75
12) 1,1-Dichloroethene	3.70	96	2592	0.956	ug/l #	79
44) Trichloroethene	8.80	130	295094	88.731	ug/l	96
64) Tetrachloroethene	10.60	164	1469	0.531	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062620\
Data File : VN062133.D
Acq On : 26 Jun 2020 12:36
Operator : JC/MD
Sample : L3043-18DL 10X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE120D1-20200615DL

Quant Time: Jun 27 04:16:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

