

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062620\
 Data File : VN062138.D
 Acq On : 26 Jun 2020 14:37
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jun 27 04:28:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	224312	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	414116	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	392266	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	150244	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	152224	48.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.72%	
35) Dibromofluoromethane	7.55	113	129375	48.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.40%	
50) Toluene-d8	10.06	98	512926	48.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.60%	
62) 4-Bromofluorobenzene	12.38	95	168337	46.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.84%	
Target Compounds						
44) Trichloroethene	8.80	130	10888	3.402	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062620\
Data File : VN062138.D
Acq On : 26 Jun 2020 14:37
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jun 27 04:28:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

