

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062620\
 Data File : VN062163.D
 Acq On : 27 Jun 2020 1:28
 Operator : JC/MD
 Sample : L3113-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DRUM-COMP

Quant Time: Jun 27 06:43:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	211451	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	410035	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	399293	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	149166	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	151407	51.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.04%	
35) Dibromofluoromethane	7.55	113	125605	47.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.50%	
50) Toluene-d8	10.06	98	516794	49.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.30%	
62) 4-Bromofluorobenzene	12.38	95	171840	48.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.74%	
Target Compounds						
16) Acetone	3.78	43	23130	21.124	ug/l	95
18) Methyl Acetate	4.29	43	4532	1.921	ug/l #	79
20) Methylene Chloride	4.50	84	9107	3.140	ug/l	96
43) Isopropyl Acetate	8.13	43	33731	6.063	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062620\
Data File : VN062163.D
Acq On : 27 Jun 2020 1:28
Operator : JC/MD
Sample : L3113-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DRUM-COMP

Quant Time: Jun 27 06:43:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

